

Quality of Service, Adaptation, and Security Provisioning in Wireless Patient Monitoring Systems

Wolfgang Leister¹, Trenton Schulz¹, Arne Lie²
Knut Grythe² and Ilangko Balasingham³

¹*Norsk Regnesentral*
²*SINTEF*

³*Interventional Centre, Oslo University Hospital*
Dept. of Electronics & Telecommunications
Norwegian University of Science & Technology
Institute of Clinical Medicine, University of Oslo
^{1,2,3}*Norway*

1. Introduction

Modern patient monitoring systems are designed to put the individual into the centre of the system architecture. In this paradigm, the patient is seen as a source of health-relevant data that are processed and transferred. Patient monitoring systems are used in health care enterprises as well as in paramedic, mobile, and home situations to foster ambient assisted living (AAL) scenarios.

There are a multitude of standards and products available to support Quality of Service (QoS) and security goals in patient monitoring systems. Yet, an architecture that supports these goals from data aggregation to data transmission and visualisation for end user has not been developed. Medical data from patient monitoring systems includes sampled values from measurements, sound, images, and video. These data often have a time-aspect where several data streams need to be synchronised. Therefore, rendering data from patient monitoring systems can be considered an advanced form of multimedia data.

We propose a framework that will fill this QoS and security gap and provide a solution that allows medical personnel better access to data and more mobility to the patients. The framework is based on MPEG-21 and wireless sensor networks. It allows for end-to-end optimisation and presentation of multimedia sensor data. The framework also addresses the QoS, adaptation and security concerns of handling this data.

In Section 2 we present background on patient monitoring systems, their requirements and how we envision communication is handled. We present communication systems in Section 3 and how to treat QoS in Section 4. A short introduction to data streaming, binary XML and how they relate to patient monitoring systems is presented in Section 5. In Section 6 we our proposed solution for the framework and present a security analysis of it in Section 7. Finally, we offer our conclusions in Section 8.



Fig. 1. Surgeons testing a patient monitoring system consisting of sensors, actuators, and communication and presentation entities.

2 Patient monitoring systems

Patient monitoring systems comprise sensors, data communication, storage, processing, and presentation of medical data. These functions are performed both near the patient, in local surgery, or remotely at a health care infrastructure, e.g., a medical centre or a hospital. An example of a patient monitoring system is shown in Fig. 1. In this figure, we see surgeons holding some sensors and actuators. On the left side of the figure, a large monitor displays the data from these devices. The information is also displayed on the laptop in the foreground of the figure.

Patient monitoring systems can be used in a variety of health care scenarios ranging from paramedic, diagnostic, surgical, post-operative, and home surveillance situations. The systems must meet a high demand of flexibility since data may be produced outside a health care enterprise. This requires specific measures in order to fulfil security, availability, privacy, and QoS demands. The properties are: *a) mobility; b) outside hospital infrastructure; c) biomedical sensor networks in use; d) wireless channel.*

As shown in a case study by Balasingham, Ihlen, Leister, Røe & Samset (2007), even within or between health care enterprises, the requirements that applications need to meet are strict and require specific measures or architectures. Data from patients are transferred through networks to the health care enterprise, and made available in a suitable form to the medical personnel to support the treatment of patients.

2.1 Communication levels

In order to account for the different health care scenarios, we propose the following *levels* surrounding the patient in which data are processed and transferred, as outlined in Fig. 2. These are divided into four levels — (0), (I), (II), and (III) — depending on the logical distance

to the patient with Level (0) being the patient. For Level (II), usually only one type applies at a time. However, it must be possible to switch between the types in Level (II) as easily as the patient moves between them.

(0) Patient. This is the actual patient.

(I) Personal sensor network. The personal sensor network denotes the patient and the sensors measuring the medical data. These sensors are connected to each other in a biomedical sensor network (BSN). While this sensor network can be connected randomly, in most cases one special BSN node—typically one that has additional power and computational resources—is appointed to be a personal cluster head (PCH), where all data for one patient are collected. The PCH may have visualisation devices attached to be used by the patient or by medical personnel. In this case, the PCH represents all data for a patient. Other topologies are possible, including the possibility that data from other patients are transferred using the sensor nodes of another patient's BSN. However, due to resource limitations of the sensor devices, organisational and ethical issues may occur. Therefore, this possibility is disregarded.

(IIa) Paramedic. In the paramedic scenario, the BSN is connected to the medical devices of an ambulance (car, plane, helicopter) via the PCH. The devices of the ambulance can work autonomously, showing the patient status locally. Alternatively, the devices of the ambulance can communicate with an external health care infrastructure, e.g., at a hospital. Note that the ambulance needs to employ some form of long-distance communication to the external health care infrastructure.

(IIb) Smart home. The smart home scenario envisages that the patient is in a smart-home environment, where the personal sensor network is connected to the infrastructure of the smart-home. The smart home infrastructure might be connected to a health care enterprise infrastructure using long-distance data communication.

(IIc) Mobility. The mobility scenario envisages that the patient is mobile, e.g., using public or personal transportation facilities. The personal sensor network of the patient is connected to the infrastructure of a health care enterprise via a mobile device, e.g., a mobile Internet connection. Note that the mobile scenario requires temporary storage in the PCH, since communication cannot be guaranteed at all times. The application and the communication software must be aware of this.

(IIId) Intensive care/surgery. During an operation the sensor data are transferred to the PCH or directly to the hospital infrastructure over a relatively short distance. The sensors are in a very controlled environment, but some sensors might be very resource limited due to their size, so extra transport nodes close to the sensors might be needed. In the operation environment, there is an increased need for QoS, so that correct data are available to the surgeons at any time during the operation.

(IIe) Pre- and postoperative. During pre- and postoperative phases of a treatment, and for use in hospital bedrooms, the sensor data are transferred from the sensor network to the PCH, and from there to the health care information system.

(III) Health care information system. The health care information system is considered a trusted environment. It comprises of the hospital network, the computing facilities, databases, and access terminals in the hospital. It should be noted that communication between Levels (II) and (III) is two-way.




Fig. 2. Generic model of patient monitoring systems showing the data flow to the observer.

Each level may have one or more data *observers*. An observer can be either the patient, medical personnel using a suitable terminal, or a processing unit that can trigger alarms, aggregate data, create logs, etc. The observer is usually in Level (II), while the communication may, or may not go through Level (III), depending on the application. The generic model in Fig. 2 helps identify where possible technology-transitions in the line of communication appear, as well as where the levels of equal security requirements can be placed.

2.2 Medical data streams

The medical data in a patient monitoring system, regardless of which level it is in, form streams of data which can be characterised as temporal multimedia data. These data streams contain the sensor data, often sampled values, attached to time-stamps and meta data, such as type and identification of the data streams. In most cases several separate streams of different data are used to describe the situation of a patient at a given time interval. The data may also contain triggers, alarms, and video data, as the capabilities of the sensor devices increase.

Multimedia data streams have a producer—here, the biomedical sensor—and a consumer—here, a (mobile) terminal or a database. In principle, each data stream can be transferred separately from the source (producer) to the sink (consumer). However, this might be impractical for improvised situations since the assurance of requirements for QoS, availability, security, and privacy will not be possible in a unified way. Therefore, we propose to forward data in a standardised way, using the system model of a generic patient monitoring system shown in Fig. 3.

This system model is suitable for patient monitoring systems where data from sensors are transferred to an observer who retrieves these data using a terminal. The medical data may be transferred to the health care information system to be stored and processed there. Additional data from the health care information system may be used by the observer at a terminal combined with the sensor data.

This generic system model divides the communication from Channels A to D as follows: Channel A includes the sensor and the sensor network to the PCH, which acts as a gateway for the personal sensor network to a network in Levels (IIa)–(IIe), also denoted as Level (II). Channel A may involve several intermediate nodes employing both wireless or wired data transfer. Channel B describes the channel from the PCH to the observer terminal, keeping the communication in Levels (II), without going through the hospital infrastructure. Channel C




Fig. 3. System model of a generic patient monitoring system, identifying the communication channels while the patient monitoring system transfers data.

transfers data from the PCH to the hospital infrastructure in Level (III) using infrastructure like the Internet, a wired carrier or a wireless carrier. Channel D describes the data transport from the hospital information system to the observer terminal in Level (II) using infrastructure like the Internet, a wired carrier or a wireless carrier.

The generic system model in Fig. 2 shows the data flow to the observer in different health care scenarios, while Fig. 3 shows the communication channels in such a system. In this architecture, note that security functions, like establishing identities for authentication, might use different channels in advance of the phase where the medical data are transferred. The different phases are presented by Leister, Fretland & Balasingham (2009).

The generic model is not dependant on how the communication in the scenarios of Level (II) are implemented. Channels C and D can have different characteristics depending on the use case, i.e., whether an external (ambulance, mobility, smart home) or an internal (hospital-related scenarios) source is used. Note also that Channel B can meet different requirements, depending on the scenario. However, from a security perspective, a short-range wireless channel is assumed.

Using the generic system model, we are able to treat the security challenges separately for every channel, thus reducing the complexity of the security analysis. However, note that each channel is implemented using several of the communication layers in OSI model.

In our framework, we intend to provide end-to-end streaming of medical sensor data as depicted in Fig. 3: *a*) from the patient to the terminal of the medic via the PCH (using Channels A and B); *b*) from the patient to the health care infrastructure (using Channels A and C); and *c*) from the health care infrastructure to the terminal of the medic (using Channel D). This also includes data streaming using Channels A, C, and D. The characteristics of these channels vary with the scenario that is addressed.

In our concept, all medical data streams are expressed using the notion of the MDI, which in some cases, e.g., in Channel A, may be expressed as μ MDI (see Section 6).

3. Communication systems

In this section, we introduce wireless sensor networks and the need for providing quality of service. We focus on the communication at Levels (I) and (II) since these are the most interesting and are in contact with the patient. At each level, we can implement a communications technology, such as ZigBee in Level (I); Bluetooth, WLAN, ZigBee or wires in

Level (IIa); Bluetooth in Level (IIc), etc. Employing only one technology in each level makes it easier to define and structure the security and QoS requirements. The medical data first must traverse Level (I), then through Level (II), and possibly arrive at the hospital infrastructure, before reaching the observer of the medical data, i.e., the medical personnel in Levels (II) or (III). It is important to have well-defined interfaces between these levels as they need to be technically implemented.

3.1 Wireless sensor networks

A wireless sensor network (WSN) is often a part of a patient monitoring system. In our reference model shown in Fig. 3, the WSN is denoted as Channel A. The WSN consists of base station receiving data from one or more tiny, low cost, low power sensor nodes that monitor information. The sensors are clustered and relay information from other sensors that may not be close enough to reach the base station.

A WSN is a good fit for a patient monitoring system since wireless technology is increasingly used in the health care industry to help eliminate cables in patient monitoring systems. Here, sensors can communicate wirelessly with a monitor that is close to the patient in a BSN. For our purposes, a BSN can be considered a special case of a WSN. The WSN can contain many small sensors that are capable of collecting vital signs and environmental information and forwarding them along to the base station; the base station can then pass this information onto the patient monitoring system. This leads, potentially, to more mobility for patients and medical staff.

While a WSN can be used in many environments and situations—for example, Römer & Mattern (2004) list applications that include herding and observing animals, checking the movement of glaciers and ocean water, and military applications—the patient monitoring systems have specific QoS and security requirements different from the other applications. For example, medical data is considered private information and wireless communication can be easily intercepted. This leads to issues in privacy, confidentiality and integrity. Also, wireless networks have their own issues with quality of service and radio interference. In addition, an attacker could alter the communication leading to threats for the patient.

The basic operations of a WSN are depicted in Fig. 4. The purpose of the deployment lies in the observation, aggregating and reporting of events in a spatio-temporal process. The communication strategies within a sensor node must support the occurrence of observed process events in various parts of the network, and possibly using distributed signal processing. Therefore, the nodes in a sensor node must cooperate to maximise the probability to fulfil their deployed mission.

There are a variety of standard solutions that are used for communication between the nodes on a wireless sensor networks. Depending upon the application and operational conditions, the most suitable is selected. ZigBee and Bluetooth are two examples out of a growing set of alternative solutions.

3.2 Quality of service provisioning

The observations made by the sensors are processed by suitable algorithms, possibly in a distributed and collaborative fashion, before the results are conveyed from the WSN to the users via a set of external networks. This way of approaching the operations of a WSN resembles the traditional encoding and transmission of a single information source (Blahut, 1987), like voice in a mobile phone. The upper reference bound for QoS experienced by the




Fig. 4. Collaboration for information generation in a WSN and its transport to an external user through the WSN and external network. S_0 is the node of origin and PCH represents the patient cluster head.

user is the entropy of the source, while the delivered QoS is a result of the capabilities and degradations induced by the chosen encoding and transmission capabilities.

An overview of QoS influential factors for sensor networks are presented and discussed by Chen & Varshney (2004). These are organised into application- and network-specific factors, with emphasis on the network. A middleware for supporting QoS in WSNs is introduced by Heinzelman, Murphy, Carvalho & Perillo (2004) with examples from medical applications. This middleware integrates the application and the WSN management to respond to the required QoS and network lifetime. A protocol-independent QoS support for WSN is presented by Troubleyn, De Poorter, Ruckebusch, Moerman & Demeester (2010) where the packets between nodes are organised according to priority processing. The integration of a WSN with external networks requires these two entities to be jointly considered (Khoshnevis & Khalaj, 2007). The external transport mechanism must be mirrored in the WSN to make QoS tradeoffs at this level. Similar aspects are also discussed by Patel & Jianfeng (2010).

Although many aspects of QoS in a WSN have been given in the literature, and solutions for networks and signal processing exist, e.g., as presented by Lei & Heinzelman (2007), a structured approach for organising and balancing the tradeoffs and degradation mechanisms appears to be lacking. In the spirit of the layered QoS for IP and mobile networks (Bai, Atiquzzaman & Lilja, 2006), we include, in the following section, the layered application QoS stack presented by Grythe, Lie & Balasingham (2009). This stack organises and separates the degradation mechanisms within a WSN for the purpose of QoS trading between the various layers. This layered approach also facilitates tradeoffs between the intra- and inter-WSN data transport.

4. Application-oriented layered QoS stack for sensor network

The purpose of the WSN deployment lies in the observation and reporting of variables or detecting events in a spatio-temporal process as indicated in Fig. 4. As such, the overall QoS to be evaluated should be oriented towards the application and end-user. The QoS experienced by the user is a result of degraded maximum theoretical information content in the event area.

The QoS degradation is due to both deployed topology and algorithmic imperfections under interaction with communication imperfections both internally and externally to the WSN. The operations of a WSN and associated systems can be split into five different actions:

- (1) Carry out the process observation by the distributed sensor nodes, each doing individual measurements.
- (2) In the case of distributed signal processing, enable the nodes to collaborate under the framework of the implemented algorithms.
- (3) Based upon the operations of the algorithms, the nodes finally reach a consensus called a result instance. This may be either periodically or a more random time domain operation.
- (4) The result instance is transmitted to the user, initiated by a random or predetermined sensor node called the node of origin, S_0 .
- (5) The result instance is presented to the user via a terminal.

4.1 Layered QoS stack

The communication strategies within a sensor node must support the occurrence of observed process events in various parts of the network and possibly distributed signal processing. Therefore, the nodes in a sensor network must cooperate to maximise the probability of meeting their deployed mission. The four-layered QoS stack of Fig. 5 simplifies the organisation of the degradation tradeoffs. Generally, the user perceptual QoS evolution between the layers behave as

$$QoS_{En} \geq QoS_{Dep} \geq QoS_{Eff} \geq QoS_{InTrans} \geq QoS_{ExTrans} = QoS_{UserInput}$$

where the QoS subscripts indicate which layer they belong to. This equation of inequalities reflects that the available user experience — or perception of QoS — is decreasing through the WSN towards the user. Examples of metrics representing the user QoS are variances, signal to noise ratio (SNR) for source coding or detection probabilities for event detections. QoS metrics are derived from the specific application, representing the most proper quality criterion. For a given WSN implementation and a given QoS metric q , the evolution of at the various levels of the layered model can be expressed as:

$$\begin{bmatrix} q_{En} \\ q_{Dep} \\ q_{Eff} \\ q_{InTrans} \\ q_{ExTrans} \end{bmatrix} = \begin{bmatrix} 1 & 0 & 0 & 0 \\ b_1 & 0 & 0 & 0 \\ 0 & b_2 & 0 & 0 \\ 0 & 0 & b_3 & 0 \\ 0 & 0 & 0 & b_4 \end{bmatrix} \begin{bmatrix} q_{En} \\ q_{Dep} \\ q_{Eff} \\ q_{InTrans} \end{bmatrix}$$

or more compact as:

$$\mathbf{q} = \mathbf{Bq}^U$$

where \mathbf{q} contains all the layer metrics while \mathbf{q}^U contains the process and intra-WSN metrics. Associating $q_0 \equiv q_{En}$ and $q_4 \equiv q_{ExTrans}$, the QoS evolution is logically expressed in a product form as:

$$q_n = q_0 \prod_{i=1}^n b_i; n = 1 : 4$$

This expression reflects the QoS interactions and tradeoff levels of Fig. 5. Due to the statistical behaviour of the influential mechanisms, as discussed later, the parameters in the matrix \mathbf{B}




Fig. 5. The application oriented layered QoS stack for sensor networks and its association with the external network.

represent random variables, $p(\mathbf{b})$; $\mathbf{b} \equiv \{b_1, b_2, b_3, b_4\}$. Their values and quantification range depends upon the chosen metrics. Defining

$$f(\mathbf{b}^n) = \prod_{i=1}^n b_i; n = 1 : 4$$

the mean value of the associated value of the quality metric is expressed within the variable domain $D(\mathbf{b}^n)$ as:

$$E(q_n|q_0) = q_0 \int_{D(\mathbf{b}^n)} f(\mathbf{b}^n)p(\mathbf{b}^n)d\mathbf{b}^n$$

As is demonstrated later, for a specific application and implemented algorithm in the WSN, it is possible that not all layers are relevant.

There are many mechanisms that influence each of the four levels. We discuss a few of them below.

Level QL0—Entropy. The Entropy level represents the maximum theoretical information content of the spatio-temporal process (STP) in the case of feature extraction (Youngchul, Poor & Heejung, 2009; Blahut, 1987) or the optimum event detection performance given the properties of the STP (Trees, 1968; Viswanathan & Varshney, 1997). These features are only related to the STP and not to any specific deployed algorithm or communication solution. As such, QL0 represents the upper performance and information bound a user can expect from the STP.

Level QL1—Deployment. This level represents the spatial sampling of the STP performed through the topology of the specific deployed WSN and the associated implemented, possibly distributed, signal processing algorithm. Given these factors, a definition of the best performance can be done assuming that all the samples from all the sensor nodes are collected and processed in a centralised unit without transport time delays or errors. This is the maximum QoS a user or users can expect from the given WSN and

does not depend upon the communication properties within or external to the WSN. The algorithm may produce a single result instance, a limited set or a continuous time discrete stream of results. In the following, we denote one quantified result as a result instance.

Level QL2—Efficiency. Efficiency denotes the QoS degradation due to interaction between the nodes in the area of event (see Fig. 4) when executing the algorithm to obtain a result instance and how this contributes to the QoS. The factors involved may be divided into topology-related and communication-influenced. Topology-related factors include node malfunction, distributed energy consumption and saving while examples of communication-influenced factors are packet loss, bit error rate (BER), variable delays and capacity. As discussed by Bai et al. (2006); Lei & Heinzelman (2007) and others, there are several factors influencing the networking within a WSN. Cross-layer optimisation and energy efficiency are key factors in this context.

Level QL3—Transport. Once the implemented algorithm produces a result instance, this is conveyed from a node of origin— S_0 in Fig. 4—to the external user via intra-WSN transport and the external network. S_0 may be a fixed node or an appointed node during the algorithmic interaction. In Fig. 5, QL3 is split into QL3I and QL3Ex to separate the internal and external transport mechanisms respectively. The typical transport degradation mechanisms are delay and packet loss in addition to broken intra links due to nonfunctional nodes. Since the interaction with the external network occurs at QL3, this layer implicitly involves the issue of optimising this interaction to balance the transport QoS degradation. In the single user case, a single external, possibly heterogeneous, network is involved; with multiple users, many different external networks are possible. The transport cost functions and packet losses are factors influencing this interaction. In some cases with small losses (Khoshnevis & Khalaj, 2007), the intra- and inter-transport may be treated separately. Otherwise, a joint optimisation is necessary involving the transport statistics of all the networks, including the intra-WSN transport.

4.2 Application examples

We give three examples of signal processing in a WSN to illustrate the involvement of the different QoS layers. We apply the QoS measure at the algorithmic level and not at an information theoretical level. The latter is reflecting the STP properties.

Single sensor and one parameter. The observation is performed with one sensor node. A time series, $\{z_m; m \in Z_+\}$, is generated and the samples are transmitted to the user, possibly via multiple hops, in the WSN. Here S_0 is also the sampling node. The reconstruction quality of the sequence by the user is governed by the QoS layers QL0, QL1, QL3I and QL3Ex. If the node is malfunctioning in this case, no information is conveyed and QL2 is influential. A proper algorithmic user QoS metric may be the minimum mean square error (MMSE):

$$\text{MMSE} \equiv E \left[(z_m - z_m^{\text{User}})^2 \right]$$

Uniform STP and sum aggregation. We assume that a WSN with K nodes is deployed in the STP area and that the area of event is the whole STP area, i.e., $A_{\text{Event}} = A_{\text{STP}}$. We want

to calculate the spatial mean value given by the aggregation function

$$E(z_m) = \frac{1}{K} \sum_{k=1}^K z_m^k = \gamma_m$$

If the PCH node is identical to the node S_0 , the layers involved are QL0, QL1, QL2 and QL3Ex since there is no intra-transport of the final result instants $\{\gamma_m\}$. Again MMSE is a proper user QoS metric.

Collaborative information processing (CIP). The nodes within A_{Event} is carrying out a CIP. In this case, a set of nodes use their observations and collaborate to generate a set of result instances at each time instance m . In other words:

$$\{Y_m\} \xRightarrow{CIP} \{\hat{X}_m\}$$

As can be seen in Fig. 4, for a deployed CIP in an event area, all the layers are contributing to the end-user QoS result. In this case, the MMSE is also applicable as an objective user QoS metric.

4.3 Local versus global optimisation

In the context of WSN for medical applications and the described variability in user terminal capabilities and heterogeneous external networks, proper QoS to each terminal is critical. Elements within a WSN influencing the QoS are briefly described by Chen & Varshney (2004). The operations of the network are governed by two modes: event reporting or queries from users. For proper operation of a medical WSN with respect to QoS and energy efficiency, external users are divided into two groups: super users and application users. Super users have access to both the PCH (Patient Cluster Head) and individual bidirectional nodes of the WSN. They can tune parameters and the performance of the WSN or initiate a deep network data query. Application users can only access the PCH and obtain streaming data from the PCH or members of parameter sets stored in the PCH. This distinction is necessary to avoid draining energy due to excessive queries and to control the access to the sensor nodes. The general description presented previously expressed the QoS evolution in a product form as:

$$q_n = q_0 \prod_{i=1}^n b_i; n = 1 : 4$$

This can be considered as an expression pointing towards a single user representation or a representation for multiple users as $U \equiv \{u_j; j = 1 : J\}$ in the situation where all the external networks are identical. If the performance of the external networks are randomly varying or are a mix of different solutions as with heterogeneous networks, the QoS evolution for each individual user $u_j \in U$ can be expressed as $q^j = b_4^j \cdot Q^0$, where the internal operations of the BAN is represented by

$$Q^0 = q_0 \prod_{i=1}^3 b_i$$

The performance trade-offs between the internal and external behaviour of the WSN to obtain the wanted or expected user QoS can be analysed and evaluated through these mechanisms. A comment that is valid for both QL2 and QL3 is that when multiple disjoint observations areas within the WSN are active simultaneously, the wireless communication activities may

create mutual interference. Also note that a similarly layered QoS structure can be used in the situation where an actuator in a WSN is fed with an excitant signal from an external user or control algorithm.

Such a layered QoS modelling resembles the data processing theorem (Blahut, 1987) stating that, in a cascade of signal processing elements, no processing can increase mutual information. The statement puts no limit on the sophistication or complexity of the individual cascaded processing blocks.

5. Data streaming

Typically, sensor data are streamed from the source to the sink, i.e., the sensor data are consumed, e.g., presented on a terminal, while new data from the same stream are arriving; an exception might be sensors that transfer single still images on demand. For data streaming, several standards have been developed covering certain properties. Relevant Internet protocol standards include the transport protocol RTP/RTCP (Schulzrinne, Casner, Frederick & Jacobson, 2003) and session protocols like SDP (Handley, Jacobson & Perkins, 2006) and RTSP (Schulzrinne, Rao & Lanphier, 1998). Relevant media compression standards include ISO MPEG 2 Systems (ISO/IEC 13818) and MPEG 4 Systems (ISO/IEC 14496). Media content searching, filtering and transactions include MPEG 7 (ISO/IEC 15938) and MPEG-21 (ISO/IEC 21000).

The DICOM standard (DICOM, 2008), which is the basis for Picture Archive and Communication Systems (PACS), is not designed to support streaming of medical data. Some research has begun to add support for streaming (Dragan & Ivetic, 2009), but this is still in the preliminary stages. On the other hand, DICOM is well suited of storing and retrieving sensor data in a health care environment. So, it should be possible to import and export to DICOM.

5.1 MPEG-21

MPEG-21 (Burnett, Pereira, de Walle & Koenen, 2006) is a standard from the International Standards Organisation (ISO) that attempts to define a complete infrastructure for delivery and consumption of multimedia content. The effort was started in 2000. The ISO's introduction to Part 1 of the standard summarises the vision as, "...to define a multimedia framework to enable transparent and augmented use of multimedia resources across a wide range of networks and devices used by different communities" (International Standards Organisation, 2004, page vii). MPEG-21 currently consists of 18 parts that cover diverse issues such as defining the fundamental structures, adapting these structures between different networks, specifying creator's and user's rights, and defining testbeds and conformance suites.

The basic unit for data in MPEG-21 is the digital item (DI). Like many other parts of MPEG-21, it is defined in XML to be machine readable. The DI is a generic item that can contain components, resources, or other containers. These structures can either refer to other XML structures, included data, or reference another item by a universal resource identifier (URI). Additional parts of MPEG-21 include how to uniquely identify an item and how to process them.

MPEG-21 defines two ways of streaming information. One is by using Binary XML that was originally part of MPEG-7, but is included as Part 16 of MPEG-21. Part 18 defines digital item streaming (DIS), which is a pure XML way of streaming.

Another goal of MPEG-21 is to ensure that creators can protect their creations and that users of content have the correct rights for viewing them. Parts 4, 5, and 6 of the standard deal with

these issues. It defines the rights—Rights Data Dictionary (RDD), how to express the rights in a machine readable form—Rights Expression Language (REL), and finally, how they are enforced—Intellectual Property Management and Protection (IPMP).

Since the vision is that MPEG-21 is to be used in a variety of situations, MPEG-21 defines Part 7—Digital Item Adaptation—for specifying how streams (i.e., the media content) should be adapted to suit different network and terminal characteristics, like bandwidth and display resolution. This is a vital mechanism in order to maximise the perceived QoS given the constraints of the network and observer equipment.

5.2 Binary XML

One of the issues with using wireless sensor networks is that the sensors cannot transfer much data in each individual packet. For example, earlier specifications of ZigBee restrict the total packet size to 128 bytes including routing information. Later versions of ZigBee allow for larger sizes, but that is dependent on the hardware in use (Daintree Networks, Inc., 2010). So, the size of the data in each packet is important. Since MPEG-21 uses XML throughout, we need to use Binary XML to compress this.

Binary XML refers to any specification which defines the compact representation of the Extensible Markup Language (XML) in a binary format (Bray, Paoli, Sperberg-McQueen, Maler, Yergeau & Cowan, 2006). While there are several competing formats, none has been accepted as a standard yet. Using a binary XML format reduces the size and eases parsing of the documents at the cost of human-readability. Binary XML is used in applications where performance or resource limitations apply. As part of this, we do not consider traditional compression methods applied to XML documents, such as using gzip, or an existing standard like ASN.1.

There are several different implementations of binary XML. For example:

- The *Fast Infoset* standard from the ISO and International Telecommunications Union (ITU) (International Standards Organisation, 2007b)
- The *Extensible XML Interchange format* from the World Wide Web Consortium (W3C) (Schneider & Kamiya, 2009; Peintner & Pericas-Geertsen, 2009)
- The *Binary MPEG format for XML (BiM)* defined by the ISO (International Standards Organisation, 2006) and is used by many ETSI standards for Digital TV and Mobile TV

There are also other binary XML formats that are defined for specific application domains.

We selected the Binary MPEG XML Format (BiM) and Efficient XML Interchange (EXI) as the most promising candidates since these were designed to work with general purpose XML, and because it was possible to get access to implementations of both.

BiM is part of the MPEG-7 standard (International Standards Organisation, 2006) and is also part of MPEG-21 (Part 16). It specifies a general method for compressing and decompressing XML documents for efficient transport and storage. It does this by examining the schemas for the document and using that information to create a separate decoder header that is used for encoding and decoding documents that use those schemas. The encoder uses this information to encode and partition the document into fragments. These fragments are then sent to the decoder. The decoder reassembles the fragments into a semantically similar document.

EXI comes from the W3C. Its purpose is to have a "...very compact, high performance XML representation that was designed to work well for a broad range of applications," (Schneider & Kamiya, 2009). It currently is a W3C Candidate Recommendation. EXI is *schema informed*, which means that it can use the schema to create a more efficient document, but a schema is not

necessary. EXI is compatible with other documents at the XML Information Set level, but not at the XML syntax level. This means that you can inter-operate with other XML documents assuming that you have everything at the Information Set level (2009). EXI does not have explicit support for fragmenting itself, instead it depends other XML technologies for this.

Ignoring the separate BiM decoding header, both EXI and BiM compress documents to comparable sizes. Therefore, other criteria were needed to figure out which method would be used. One obvious advantage that BiM has over EXI is its support for fragmentation and streaming. The BiM standard (International Standards Organisation, 2006) goes into detail about how BiM can be fragmented and reassembled, while EXI (Schneider & Kamiya, 2009) only describes that EXI fragments can exist, but not how a document could be fragmented, streamed or reassembled.

6. Proposed solution

We have experimented with a variety of different solutions and have come up with some solutions that integrate both WSN and MPEG-21 into patient monitoring systems. As the basic idea, we adapt MPEG-21 to a WSN by adapting the basic Digital Items and employing binary XML.

6.1 MPEG-21 for patient monitoring systems

MPEG-21 includes features like adaptation and security for network and users that make it a good candidate in a health care system architecture. However, most of the patient data aggregators are based on wireless sensor networks containing nodes with restricted capabilities and battery capacity. This means that they may not be able to run all of the processor intensive tasks of MPEG-21. Therefore, the MPEG-21 standard needs to be adapted to work with this restricted setup.

As mentioned in Section 5.1, a digital item (DI) is the basic unit for transferring information. Landén (2003) introduces a medical digital item (MDI) that adapts the digital item to the health care sector. The MDI also encapsulates mechanisms for adaptation and security. Since sensors have limited resources, we developed a *lightweight MDI* or μ MDI for the sensors. However, to save battery capacity and processing power in some situations, the XML syntax of MPEG-21 is not employed at all.

Lie, Grythe & Balasingham (2008) present a hospital scenario where the use of digital item adaptation (DIA) is designed for sensors wirelessly connected to a PCH. The PCH is connected to the health information system. The medical personnel access the patient data using a terminal with an MPEG-21 enabled browser. After a threat analysis of patient monitoring systems (Leister, Abie, Groven, Fretland & Balasingham, 2008) we extended this architecture to include (BSNs) (Leister, Fretland & Balasingham, 2008). The architecture supports security, privacy, and authentication, by using the mechanisms of MPEG-21 in the OSI presentation layer and a carefully designed μ MDI (Leister et al., 2009). It also supports content adaptation according to terminal and network capabilities, QoS, energy efficiency, and several types of session mobility. The impact of this work is currently being studied in a recently developed testbed for evaluating MPEG-21 in health care applications, which is described in Section 6.4. Other approaches to using MPEG-21 in health care use the IPMP Part of MPEG-21 for patient records (Brox, 2005). Recently, a framework for using MPEG-21 IPMP components for a security framework for pervasive health care architectures has been presented by Fragopoulos, Gialelis & Serpanos (2009), including wireless communication between personal

digital assistants (PDA). While that work introduces MPEG-21 for medical applications, our work includes the use of MPEG-21 for BSN and uses a generic model to analyse the threats.

6.2 Using binary XML

The initial plan was to use BiM since we could partition up a document into fragments, compress it, and send it all with one tool. EXI would only give us the compression. Ultimately, the final decision came down to the quality of implementations that were available since we did not have time nor resources to create our own implementation. The BiM implementation (Technische Universität München, 2007-2008) seems to be incomplete, is not documented, and has many bugs. Initially, it would not encode our μ MDI document at all. After a lot of tracing and fixing, we were able to get it to encode a document similar to μ MDI, but it then failed to decode it correctly. It also seems to be missing the code for partitioning the document into fragments and reassembling it.

By contrast, the EXI implementation (EXIficient) is available as an open source project (Peintner, 2010). While it is missing some documentation, it is actively developed and seems to handle all the situations we have given it. Since EXI has no built in streaming, we needed a method for streaming the items. This was solved using MPEG-21 Part 18—Digital Item Streaming.

Digital Item Streaming (DIS) works by specifying a separate XML document that defines the Bitstream Binding Language (BBL) for a document. This BBL specifies the packets or packet streams for the document, when these packets should be sent, and how the source document should be divided up. The BBL is then used by software to divide the document up and send it over arbitrary protocols. DIS also allows specifying constraints for packet streams (such as number of items, or packet size), but the standard indicates that constraints *will* be broken in order to send a packet. There is a publicly available implementation from the ISO (International Standards Organisation, 2007a) that works well. One issue with DIS is that it only works with with a “completed” document. That is, one cannot add siblings to the document after you have sent over the basic skeleton. Future packets must be children of the current elements.

6.3 The complete system

We combine μ MDI, EXI, and DIS to form the following solution for a BSN using MPEG-21. We have expanded the μ MDI to contain multiple entries. A portion of the new μ MDI can be seen in Fig. 6. The various sensors can then send their identification information and their data as separate packets using DIS. These packets are eventually received by the base station and are attached to an already existing skeleton document. This document has a limited number of parts. After this document is done, a new skeleton is added and the sensor can then retransmit its identity and continue sending its information. The base station can transfer this information on to the patient monitoring system.

The resulting packets should be small, but we can further reduce their size by using EXI and transmitting them. The base station will receive the items as an EXI stream and will apply them to the XML document.

The packets are sent from the WSN to a PCH. The PCH takes the information and converts it from a μ MDI to a real MDI, adds extra information (e.g., identifying information for the patient), and sends the information to other communication levels. This information is encrypted with strong encryption, ensuring that only medical staff with correct permissions may use it.

```

<did:DIDL xmlns:did="urn:mpeg:mpeg21:2002:02-DIDL-NS"
  xmlns:xsi="http://www.w3.org/2001/XMLSchema-instance"
  xmlns:dii="urn:mpeg:mpeg21:2002:01-DII-NS"
  xsi:schemaLocation="urn:mpeg:mpeg21:2002:02-DIDL-NS schemas/didl.xsd
    urn:mpeg:mpeg21:2002:01-DII-NS schemas/dii.xsd">
  <did:Container>
    <did:Item id="myitem">
      <did:Descriptor>
        <did:Statement mimeType="text/xml">
          <dii:Identifier>urn:grid:a1-abcde-9873216540-f</dii:Identifier>
        </did:Statement>
      </did:Descriptor>
      <did:Descriptor>
        <did:Statement mimeType="text/xml">
          <dii:Type>urn:sensor:bloodpressure</dii:Type>
        </did:Statement>
      </did:Descriptor>
      <did:Item>
        <did:Descriptor>
          <did:Statement mimeType="text/plain">20</did:Statement>
        </did:Descriptor>
        <did:Component>
          <did:Resource mimeType="text/plain">68.300</did:Resource>
        </did:Component>
      </did:Item>
      <did:Item>

```

Fig. 6. Excerpt from a μ MDI document adapted for Digital Item Streaming (DIS)

6.4 MPEG-21 testbed

A proposed architecture's performance and functionality is best verified using a dedicated software testing platform. Thus, a Java-based testbed was created to test the hospital and home-care networks. The testbed is based on an existing MPEG-21 software that was developed during the MPEG-21 standardisation work (Perkis & Drege, 2008), and then modified to support the aforementioned medical network requirements (Lie et al., 2008). The testbed design requirements had the following architecture in mind:




Fig. 7. The hospital network architecture envisioned for the MPEG-21 supporting testbed (Lie et al., 2008).

- Support for wireless medical sensors
- PCH bed-side terminals (optionally compatible with the NCAP—Network Capable Application Processors—architecture of the IEEE 1451 standard) (IEEE1451.0, 2007; IEEE1451.5, 2007), including WSN radio interface and WLAN radio interface
- WLAN APs and backbone network infrastructure
- User terminals equipped with MPEG-21 browsing capabilities
- An MPEG-21 Broker server (MB)
- One or several media servers (MS)
- Optional Media resource Adaptation Server (MAS)

The envisioned system is depicted in Fig. 7. The medical body sensors are communicating with the bed-side PCH terminal. This PCH includes NCAP functionality in case IEEE 1451.5 is in use, and uses short-range radio systems to communicate with the active sensors (e.g., Zigbee). Thus, the PCH will collect all the necessary meta information regarding the sensor capabilities and current status collected in MPEG-21 validated XML DIs, based on the IEEE 1451 TEDS (Transducer Electronic Data Sheets). In addition, there must be a dedicated authentication procedure to ensure that new sensors are connected to the intended PCH (the short-range radio coverage can reach other patients sensors and PCH). The principle of the *resurrecting duckling* (Stajano & Anderson, 1999) with physical imprinting between the sensor nodes and correct PCH is a viable authentication approach for exchanging symmetric secret keys.

All PCH terminals are connected to the network infrastructure, e.g., using WLAN. This helps minimise the traffic from the energy constrained sensors. A moving patient without a PCH would call for frequent handovers between the sensors and different access points.

To facilitate easy retrieval of all information available, a few centralised servers are needed. The media broker (MB) server will hold the MPEG-21 CDI (Content Description Items) that gives the updated menus on the end-users terminals, e.g., regarding patient lists. Patient and diagnosis search could be facilitated through the accompanying MPEG-7 standard and underlying meta information. Each patient will have one dedicated CDI menu system for its PCH and one for the Electronic Patient Record (EPR) system, holding all relevant EPR and pre-stored medical media. When pre-stored media is selected, it will be retrieved from the media streaming server (MS). An optional powerful adaptation server (MAS) can be used to




Fig. 8. The Java based MPEG-21 testbed architecture.

facilitate media format adaptation on-the-fly for both live and pre-stored media, following MPEG-21 Part 7 methodology (DIA — Digital Item Adaptation).

The observer will thus use the system via the terminal, e.g., an MPEG-21 capable browsing terminal, to access both live content streamed via the PCH, and pre-stored content from the MS. The MPEG-21 testbed also incorporates support for Session Mobility (Soma, 2007). The testbed is programmed in Java and ECMAScript, and has MPEG-21 validated XML files that format the graphical user interface menus with the supported functionality. In addition to supporting session mobility and DIA, it uses the MPEG-21 DIP — Digital Item Processing — engine to process the user activities. For example, to process media streaming it activates a Java based player named *Fobs4JMF*, which consists of Sun Java Media Framework (JMF) player enhanced with codecs from the open source package *ffmpeg* (Omnividea, 2008). Fig. 8 shows a diagram with the key software components. At top, the user terminal is shown. The initial GUI is formed by the local root DID (`smdid.xml` in this example). From this GUI the user can select the patient list CDI located at the MB. After selecting a specific patient, the user can choose either PCH or EPR. If choosing the PCH, the terminal will be given access over HTTP to the online PCH for this patient. This PCH contains the current CDI showing a list of connected wireless sensors (`cdi_pch.xml`). The user will now be able to access online data such as ECG. Available media format choices are stored in dedicated XML files (the




Fig. 9. Snapshots from the MPEG-21 testbed, observer terminal part: *a)* start screen with selection of source, *b)* The GUI is generated given the CDI contents, where the observer can select the particular content, and *c)* the rendering of the selected medical content.

oval object at figure mid bottom). The available media formats are checked against the UED (Usage Environment Descriptions) information, and a DIA engine located in the PCH or at a centralised MS or MAS would eventually adapt the content for optimal fidelity to the user's terminal. Finally a message will be sent to the terminal containing the URL to the possibly adapted media stream.

In Fig. 9, an example walkthrough of the user terminal functionality is displayed. The user selects a specific PCH, using IP address and path to CDI. The list of medical content described in this particular CDI is used to generate the GUI presented to the user. Behind the scenes, the selected content is adapted to the current display capabilities and network conditions, using the DIA functionality on the UED information.

7. Security analysis

In this section, we perform a security analysis of the use of MPEG-21 in patient monitoring systems. The security requirements for patient monitoring systems are given by Leister et al. (2009) as the following: *(a)* availability, *(b)* data confidentiality, *(c)* data integrity, and *(d)* data authenticity. The general threats and attacks to these security objectives are eavesdropping, denial of service, masquerading, and disclosure. Leister et al. identified the threats associated to these factors as compromised, destroyed, malfunctioning, lost or stolen components; software errors; misuse of emergency access; and denial of service attacks for the components; compromised or unstable communication infrastructure; and eavesdropping on the communication channels.

We consider these security goals separately for each of the levels introduced earlier in Section 2.1. This allows us to consider separately the threats and remedies for security, and also consider whether the particular component or channel of a patient monitoring system is within a trusted environment. Components and channels in a trusted environment can be considered secure. For components and channels outside a trusted environment, we consider the threat model by Dolev & Yao (1981), where an attacker can overhear, intercept, and synthesise any message, and is only limited by the constraints of the cryptographic methods used. Since some of the devices used may not have the resources for strong cryptographic methods, the assumption of the Dolev-Yao attacker might be not applicable in all situations,

and additional measures might be necessary, especially when wireless sensor networks are a part of the patient monitoring system.

Using the Dolev-Yao threat model, availability cannot always be guaranteed since an attacker always can synthesise messages, and thus disturb or interrupt a channel. While this cannot be prevented on the OSI application layer using MPEG-21, there might be other technical measures on lower OSI layers that can be applied, such as spread spectrum technologies. The development of radio access methods based on time-hopping and frequency-hopping spread-spectrum (TH/FH-SS) addresses power-efficient signaling methods, which becomes an interesting candidate for wireless sensor networks. TH/FH-SS is perceived as an anti-jamming technique with low complexity and does not require complicated signal processing. An example of their implementation can be found in Bluetooth systems (Tsai & Chang, 1994; Salonidis et al., 2005). However, these technologies are beyond the scope of this chapter.

Some threats and vulnerabilities can be addressed on an administrative level, such as describing responsibilities for the stakeholders, educating personnel, or providing routines for keeping devices out of the hands of adversaries. These non-technical routines are not considered further in this context.

Due to the nature of the biomedical data, we consider a model where a numbered sequence of unique data packets that are on a channel are transferred from source to sink, protected by means of MPEG-21. When this data stream arrives, packets may have been lost, altered, replayed, be out of order, be subject to eavesdropping, be stored, etc. These characteristics may be the result of processes at lower OSI layers.

In patient monitoring systems, data are usually sent in packets using both connection-based or connectionless channels. Connectionless channels do not offer retransmission automatically. However, mechanisms for retransmissions on any channel might block sending of data, and thus prohibit the timely transfer of data. Ideally, short, self-contained sequences of packets are preferred to avoid the issues in retransmission.

For Level (0) the security considerations are limited to human communication patterns, such as outlined in the communication theory by Watzlawick, Beavin & Jackson (1967). The other levels can be considered more from a technical perspective, while both technical and administrative means can be used to reduce the vulnerabilities.

7.1 Security considerations for level (I)

Level (I) is usually implemented as a wired or wireless, biomedical sensor network, where most of the data traffic consists of a stream from the sensor node to the sink since we are not considering actuator networks here. Leister et al. (2009) discuss threats for biomedical sensor networks in the domains of the entity domain, the network domain, the specific protocol domain, and the routing and forwarding domain. In these domains, an adversary who has access to the medium, can eavesdrop, inject, delete and alter data packets. This can lead to incorrect data, lost data, data being attached to the wrong data stream or patient, and inferior service quality.

Eavesdropping obviously threatens confidentiality and privacy goals. Lost, delayed, and altered packets threaten the integrity and the authenticity of the data as well as the availability. As a consequence, this could cause inappropriate, delayed or missing treatment of the patient at the risk of damage to the patient's health.

MPEG-21 does not solve threats arising at the lower communication layers, but is suited to detect possible inconsistencies in the case of altered content so that wrong content can be

marked or discarded. Regarding eavesdropping, MPEG-21 offers mechanisms to encrypt the content. Digital Items can be identified in MPEG-21, thus facilitating that the data belong to the correct stream and patient as long as the data source can be authenticated. Authentication is done by using the identification tags on the data and encryption. Therefore, as long as the encryption key is secret, the authentication will work. The initialisation procedure of sensor nodes shown by Leister et al. (2009) guarantees this. Here, the sensor nodes are provided with the necessary security credentials in the initialisation step, while these are used in the operation mode.

In wireless sensor networks, the data streams from the sensor to the PCH can be sent via other nodes. These nodes do not possess the capabilities to decrypt the streams from another sensor node, and cannot inject valid packets, or alter packets consistently. However, data aggregation in the sensor network is not possible in such a scenario.

Due to resource constraints in the devices, the possible cryptographic methods that can be applied are weak. Therefore, additional measures must be taken, such as restricting all data that can identify a patient in the wireless sensor network. The identifying information necessary to attach the data to the right patient must therefore be formed such that the patient is not revealed.

The architecture by Leister et al. suggests that the software loaded into the sensor nodes is tied together with the necessary credentials for encryption and authentication. Therefore, once provided with the credentials, the software in the sensor node cannot be altered unless exchanged entirely, which would be similar to an attacker providing a fake node.

7.2 Security considerations for level (II)

Compared to Level (I), there are less resource restrictions for the scenarios on Level (II). Therefore, better cryptographic methods can be applied, and the restriction on not sending identifying information is less important. The health care infrastructure can perform data aggregation if necessary, and also convert the μ MDI to an MDI with data not available in the sensor network. Note, however, that identifying information should not be sent in the channel unless necessary. For other characteristics, the same mechanisms apply regarding the use of MPEG-21.

When roaming between the Levels (IIa) to (IIe) the authentication mechanism beyond MPEG-21 is responsible for establishing the trust to facilitate the communication between the PCH and the entity responsible for the respective level or scenario by using the correct keys. Note that in some scenarios the PCH can have the role of both.

7.3 Security considerations for level (III)

Devices that are in the hospital infrastructure and that can be authenticated are considered secure, and not tampered with; for these devices, no specific measures are necessary. However, note that authentication is always necessary. In Level (III), the use of MPEG-21 provides a unified method to represent the medical data, and to provide data confidentiality, data integrity, and data authenticity through the mechanisms used.

7.4 Security considerations end-to-end

While MPEG-21 can be used to provide the security goals from end-to-end, other standards might be employed in a health care enterprise, such as DICOM, ELIN XML (Christensen, 2009) or HL7 (Shaver, 2007), and make data conversions necessary for the data to be stored or

processed. The representation of medical data in a MDI data structure provides the security goals of confidentiality, data integrity, and data authenticity.

Since being placed in the presentation layer, MPEG-21 can be used in the Levels (I)-(III) and end-to-end to sustain the security goals. Functionalities of the presented architecture, such as converting the μ MDI to MDI, are not necessarily performed by the PCH, i.e., Level (I), but can be performed in Level (II) or (III), before being further processed and stored. The end-to-end security is sustained as long as the entities processing MPEG-21 can be trusted, and the cryptographic methods used are strong enough.

The architecture shown here addresses the threats of devices being tampered with, eavesdropping, altered communication at several layers, in the sense that wrong data are not accepted. The limitations are in the relatively weak encryption in the sensor nodes. An adversary could eavesdrop on the traffic, break the encryption, and either introduce a fake node, or use an attack at the wireless network level. This is not considered by the Dolev-Yao attacker, but could be done in a practical case.

However, since the wireless communication using weak encryption is limited to Level (I) where the communication range is limited, an adversary needs to have access to the vicinity of the patient, or have access to communication credentials from the inside. Such threats need to be addressed on an administrative and stakeholder level, e.g., by strict control of access to areas where security is critical, rules for handling devices, measures to increase awareness among medical personnel, and so on.

The Dolev-Yao attacker sets requirements on the encryption. However, this is problematic due to limited resources in the sensors. Therefore, to avoid the risk that an attacker can get the data to the right patient, we do not include data into the stream that can be used to identify the patient, such as a personal identification number. The identity establishment must be done by routines using the mechanisms described by MPEG-21.

The security analysis shows that the communication in patient monitoring systems can be secure in principle in relation to the mentioned security goals and threats. For a concrete implementation, a formal analysis is necessary, where the involved protocols are modelled, and formally analysed using tools such as those shown by Hagaliletto, Strand, Leister & Groven (2009) in another application area. The same methods can be applied for patient monitoring systems, using concrete implementations and protocols.

8. Conclusion

We have shown a framework that includes QoS, adaptation, and security provisioning for wireless patient monitoring systems with an emphasis on a solution that integrates the patient monitoring system into the hospital infrastructure. For this, we proposed an architecture that uses the MPEG-21 standard to reach these goals. We implemented parts of the system in a testbed to show that the architecture is viable, and gained practical results.

We found that MPEG-21 is suitable for integrating a biomedical WSN with the health care infrastructure in different health care scenarios. Challenges arising from the scarce resources generally available in WSN are addressed by utilising XML compression, XML streaming, and architectural means that avoid handling XML data in its text form in the sensor nodes. One issue that remains is what should be part of the metadata in a digital item (DI) and what should be implemented in the DI adaptation engine.

We showed that security concerns can be covered by describing the scenarios and considering the different communication levels and communication channels. However, the WSN has scarce resources that result in weak cryptography and the communication medium is

accessible; the result is that not all threats can be stopped at the OSI presentation level. Therefore, in the case of certain attacks by a man in the middle, the availability might be reduced. Adaptation for a patient monitoring system is shown by the means of a testbed that also shows the integration with the information system of the healthcare enterprise.

While MPEG-21 covers security and adaptation issues, the QoS requirements are addressed in WSN by a stack-oriented approach using a degradation model in an optimisation process. How to implement these QoS structures in MPEG-21 is not investigated. However, it appears that the adaptation engine approach is more versatile than editing metadata, and permits a wider range of dynamic QoS attributes to be supported.

Future work on the architecture includes more integration and implementation on a larger scale to determine how we can adjust parameters so the architecture is suitable for a real network. These considerations about the architecture and its parameters are important for factors like power—thus battery life of sensors—achievable security, availability, and flexibility in the form of adaptation.

In the future, this work can lead to benefits for both the health care workers and patients. Using our architecture, patients are able to move around more freely, stay productive out of a hospital environment and know that their medical data is secure. At the same time, health care workers have a better overview of a patient's medical data in a variety of environments and are assured of a certain level of QoS. The result is that, in the eventuality that a patient needs it, treatment can arrive in a timely manner.

9. Acknowledgements

This research is funded by the SAMPOS and WISENET projects in the VERDIKT framework of the Norwegian Research Council.


10. References

- Bai, H., Atiquzzaman, M. & Lilja, D. J. (2006). Layered view of QoS issues in IP-based mobile wireless networks, *Int. Journal on Comm. Systems* 18: 141–161.
- Balasingham, I., Ihlen, H., Leister, W., Røe, P. & Samset, E. (2007). Communication of medical images, text, and messages in inter-enterprise systems: A case study in Norway, *IEEE Transactions on Information Technology in Biomedicine* 11(1): 7–13.
- Blahut, R. E. (1987). *Principles and practice of information theory*, Addison-Wesley Longman Publishing Co., Inc., Boston, MA, USA.
- Bray, T., Paoli, J., Sperberg-McQueen, C. M., Maler, E., Yergeau, F. & Cowan, J. (2006). Extensible Markup Language (XML) 1.1 (second edition), World Wide Web Consortium, Recommendation REC-xml11-20060816.
- Brox, G. A. (2005). MPEG-21 as an access control tool for the national health service care records service, *Journal of Telemedicine and Telecare* 11 Suppl 1: 23–5.
- Burnett, I. S., Pereira, F., de Walle, R. V. & Koenen, R. (eds) (2006). *The MPEG-21 Book*, John Wiley & Sons. ISBN 0-47001011-8.
- Chen, D. & Varshney, P. K. (2004). QoS support in wireless sensor networks: A survey, *2004 International Conference on Wireless Networks (ICWN 2004)*.
- Christensen, T. (2009). Sluttrapport ELIN, *Sluttrapport*, Den norske legeforening, Pb. 7000 St. Olavs plass, NO-0130 Oslo, Norway. In Norwegian.
- Daintree Networks, Inc. (2010). Zigbee specification comparison matrix. Last accessed September 2, 2010, <http://www.daintree.net/resources/spec-matrix.php>.

- Digital Imaging and Communications in Medicine (DICOM)* (2008). <http://medical.nema.org/dicom/>. Last accessed May 15, 2009.
- Dolev, D. & Yao, A. (1981). On the security of public key protocols, *Proc. of the IEEE 22nd Annual Symposium on Foundations of Computer Science*, pp. 350–357.
- Dragan, D. & Ivetic, D. (2009). *An Approach to DICOM Extension for Medical Image Streaming*, DAAAM International, chapter 4 in DAAAM International Scientific Book 2009, pp. 25–34.
- Fragopoulos, A., Gialelis, J. & Serpanos, D. (2009). Security framework for pervasive healthcare architectures utilizing MPEG-21 IPMP components, *International Journal of Telemedicine and Applications* 2009: 1–9.
- Grythe, K., Lie, A. & Balasingham, I. (2009). Interactive system design and end-to-end optimization in sensor network, *IEEE 5th Int. Workshop on Wearable Micro and Nano Technologies for Personalized Health*, Oslo.
- Hagalissetto, A. M., Strand, L., Leister, W. & Groven, A.-K. (2009). Analysing protocol implementations, in F. Bao, H. Li & G. Wang (eds), *ISPEC*, Vol. 5451 of *Lecture Notes in Computer Science*, Springer, pp. 171–182.
- Handley, M., Jacobson, V. & Perkins, C. (2006). SDP: Session Description Protocol, RFC 4566 (Proposed Standard), <http://www.ietf.org/rfc/rfc4566.txt>.
- Heinzelman, W. B., Murphy, A. L., Carvalho, H. S. & Perillo, M. A. (2004). Middleware to support sensor network applications, *Network*, *IEEE* 18(1): 6–14.
- IEEE1451.0 (2007). IEEE standard for a smart transducer interface for sensors and actuators - common functions, communication protocols, and transducer electronic data sheet (TEDS) formats, *IEEE Std 1451.0-2007* pp. 1–335.
- IEEE1451.5 (2007). IEEE standard for a smart transducer interface for sensors and actuators wireless communication protocols and transducer electronic data sheet (TEDS) formats, *IEEE Std 1451.5-2007* pp. C1–236.
- International Standards Organisation (2004). Information technology—multimedia framework (MPEG-21—part 1: Vision, technologies and strategy, *Technical Report ISO/IEC TR 21000-1*.
- International Standards Organisation (2006). Information technology—MPEG system technologies—part 1: Binary MPEG format for XML, *Technical Report ISO/IEC 23001-1*.
- International Standards Organisation (2007a). Directory listing for /publiclyavailablestandards /iso.iec.21000-8.2008.amd.1.2009.reference.software/. Last accessed September 8, 2010, http://standards.iso.org/ittf/PubliclyAvailableStandards/IS_IEC_21000-8_2008_Amd.1_2009_Reference_Software/.
- International Standards Organisation (2007b). Information technology—generic application of ASN.1: Fast Infoset, *Technical Report ISO/IEC 24824-1: 2007 / ITU-T Rec X891*.
- Khoshnevis, B. & Khalaj, B. (2007). Optimum QoS classes in interworking of next generation networks, *Communications and networks* 9(4 (Dec)): 438–445.
- Landén, M. (2003). *An MPEG-21 approach to creating the first multimedia electronic patient journal system*, Master's thesis, NTNU.
- Lei, C. & Heinzelman, W. B. (2007). A survey of routing protocols that support QoS in mobile ad hoc networks, *Network*, *IEEE* 21(6): 30–38.
- Leister, W., Abie, H., Groven, A.-K., Fretland, T. & Balasingham, I. (2008). Threat assessment of wireless patient monitoring systems, *Proc. ICTTA'08*, Damascus, Syria.
- Leister, W., Fretland, T. & Balasingham, I. (2008). Use of MPEG-21 for security

- and authentication in biomedical sensor networks, *Proc. ICSNC'08, International Conference on Systems and Networks Communication* 0: 151–156.
- Leister, W., Fretland, T. & Balasingham, I. (2009). Security and authentication architecture using MPEG-21 for wireless patient monitoring systems, *International Journal on Advances in Security* 2(1): 16–29, <http://www.iariajournals.org/security/>.
- Lie, A., Grythe, K. & Balasingham, I. (2008). On the use of the MPEG-21 framework in medical wireless sensor networks, *Proc. of the IEEE 1st Int Symp on Applied Sciences in Biomedical and Communication Technologies (ISABEL)*, Aalborg, Denmark, 25–28. October.
- Omnividea, F. (2008). Object oriented wrapper for ffmpeg, <http://fobs.sourceforge.net/>. Last accessed August 23, 2010.
- Patel, M. & Jianfeng, W. (2010). Applications, challenges, and prospective in emerging body area networking technologies, *Wireless Communications, IEEE* 17(1): 80–88.
- Peintner, D. (2010). EXIFcient—Open Source Implementation of the W3C Efficient XML Interchange (EXI) format. Last accessed September 8, 2010, <http://exifcient.sourceforge.net>.
- Peintner, D. & Pericas-Geertsen, S. (2009). Efficient XML interchange (EXI) primer, W3C *working draft*, W3C. <http://www.w3.org/TR/2009/WD-exi-primer-20091208/>.
- Perkis, A. & Drege, P. (2008). MPEG-21 test-bed (joadjo).
- Römer, K. & Mattern, F. (2004). The design space of wireless sensor networks, *IEEE Wireless Communications* 11(6): 54–61.
- Salonidis, T., Bhagwat, P., Tassiulas, L. & LaMaire, R. (2005). Distributed topology construction of bluetooth wireless personal area networks, *Selected Areas in Communications, IEEE Journal on* 23(3): 633–643.
- Schneider, J. & Kamiya, T. (2009). Efficient XML interchange (EXI) format 1.0, *Candidate recommendation*, W3C. <http://www.w3.org/TR/2009/CR-exi-20091208/>.
- Schulzrinne, H., Casner, S., Frederick, R. & Jacobson, V. (2003). RTP: A Transport Protocol for Real-Time Applications, RFC 3550 (Standard), <http://www.ietf.org/rfc/rfc3550.txt>.
- Schulzrinne, H., Rao, A. & Lanphier, R. (1998). Real Time Streaming Protocol (RTSP), RFC 2326 (Proposed Standard), <http://www.ietf.org/rfc/rfc2326.txt>.
- Shaver, D. (2007). HL7 101: A Beginner's Guide, *For the Record* 19(1): 22.
- Soma, C. W. (2007). Transparent consumption of digital items between terminals using mpeg-21 session mobility, *Technical report*, Norwegian University of Science and Technology. Master thesis.
- Stajano, F. & Anderson, R. (1999). The resurrecting duckling: Security issues for ad-hoc wireless networks, *Security Protocols, 7th International Workshop Proceedings, Lecture Notes in Computer Science*.
- Technische Universität München (2007-2008). MPEG-7 reference SW. Last accessed September 8, 2010, <http://www.lis.ei.tum.de/index.php?id=131>.
- Trees, H. L. V. (1968). *Detection, estimation and modulation theory*, John Wiley and Sons.
- Troubleyn, E., De Poorter, E., Ruckebusch, P., Moerman, I. & Demeester, P. (2010). Supporting protocol-independent adaptive QoS in wireless sensor networks, *Sensor Networks, Ubiquitous, and Trustworthy Computing (SUTC), 2010 IEEE International Conference on*, pp. 253–260.
- Tsai, Y.-R. & Chang, J.-F. (1994). Using frequency hopping spread spectrum technique to combat multipath interference in a multiaccessing environment, *Vehicular Technology, IEEE Transactions on* 43(2): 211–222.

- Viswanathan, R. & Varshney, P. K. (1997). Distributed detection with multiple sensors, I. fundamentals, *Proceedings of the IEEE* 85(1): 54–63.
- Watzlawick, P., Beavin, J. H. & Jackson, M. D. D. (1967). *Pragmatics of Human Communication: A Study of Interactional Patterns, Pathologies, and Paradoxes*, W. W. Norton & Company.
- Youngchul, S., Poor, H. V. & Heejung, Y. (2009). How much information can one get from a wireless sensor network over a correlated random field?, *Information Theory, IEEE Transactions on* 55(6): 2827–2847.



Biomedical Engineering, Trends in Electronics, Communications and Software

Edited by Mr Anthony Laskovski

ISBN 978-953-307-475-7

Hard cover, 736 pages

Publisher InTech

Published online 08, January, 2011

Published in print edition January, 2011

Rapid technological developments in the last century have brought the field of biomedical engineering into a totally new realm. Breakthroughs in materials science, imaging, electronics and, more recently, the information age have improved our understanding of the human body. As a result, the field of biomedical engineering is thriving, with innovations that aim to improve the quality and reduce the cost of medical care. This book is the first in a series of three that will present recent trends in biomedical engineering, with a particular focus on applications in electronics and communications. More specifically: wireless monitoring, sensors, medical imaging and the management of medical information are covered, among other subjects.

How to reference

In order to correctly reference this scholarly work, feel free to copy and paste the following:

Wolfgang Leister, Trenton Schulz, Arne Lie, Knut Grythe and Ilanko Balasingham (2011). Quality of Service, Adaptation, and Security Provisioning in Wireless Patient Monitoring Systems, Biomedical Engineering, Trends in Electronics, Communications and Software, Mr Anthony Laskovski (Ed.), ISBN: 978-953-307-475-7, InTech, Available from: <http://www.intechopen.com/books/biomedical-engineering-trends-in-electronics-communications-and-software/quality-of-service-adaptation-and-security-provisioning-in-wireless-patient-monitoring-systems>

INTECH

open science | open minds

InTech Europe

University Campus STeP Ri
Slavka Krautzeka 83/A
51000 Rijeka, Croatia
Phone: +385 (51) 770 447
Fax: +385 (51) 686 166
www.intechopen.com

InTech China

Unit 405, Office Block, Hotel Equatorial Shanghai
No.65, Yan An Road (West), Shanghai, 200040, China
中国上海市延安西路65号上海国际贵都大饭店办公楼405单元
Phone: +86-21-62489820
Fax: +86-21-62489821

© 2011 The Author(s). Licensee IntechOpen. This chapter is distributed under the terms of the [Creative Commons Attribution-NonCommercial-ShareAlike-3.0 License](#), which permits use, distribution and reproduction for non-commercial purposes, provided the original is properly cited and derivative works building on this content are distributed under the same license.