Electric and Nonelectric Quantities Measurement in Power Engineering using LabVIEW

Jan Machacek¹, Jan Slezinger² and Jiri Drapela³
Brno University of Technology
Czech Republic

1. Introduction

We need to measure both electric and nonelectric quantities solving the series of problems in power engineering. Graphical programming environment of the program LabVIEW seems to be useful helper and provides good tool for measuring both an automatic independent measurements of specific tasks, using measuring cards and instruments. The LabVIEW belongs to widely used measuring tools at Department of Electrical Power Engineering, Faculty of Electrical Engineering and Communication power engineering, Brno University of Technology. Since the LabVIEW is used to solve a number of projects at our department and there is only limited space in this article, we decided to mention only the most interesting ones. The first project deals with measuring operating characteristics of the photovoltaic cell. It consists of standard interconnection of measuring devices (producer Metex), which measure data that are transferred by program LabVIEW to the computer. Next, the data are mathematically and graphically processed. Different projects in LabVIEW offer a wide spectrum of front panels in the framework of this program allowing to connect each measuring device by its specific way. Some of them are listed here.

Another of the many tasks developed at our department is an autonomous measurement of electric and non-electric quantities of photovoltaic and wind power installation. A specific part of this project included the connection of meteorological station VIASALA which required a particular set of control interfaces created by LabVIEW. The LabVIEW environment has also created a comprehensive measurement system for detection of resistance of modern light sources to interference in the voltage supply, as well as a system to test and develop devices to measure the flickering of light sources. The system is implemented on the PXI platform.

First describes control of serial communication in LabVIEW software here, because serial communication is a popular means of transmitting data between a computer and a peripheral device such as a programmable instrument or even another computer. Serial communication uses a transmitter to send data, one bit at a time, over a single communication line to a receiver. You can use this method when data transfer rates are low or you must transfer data over long distances. Serial communication is popular because most computers have one or more serial ports, so no extra hardware is needed other than a cable to connect your instrument to the computer (or two computers together). In many applications, the electronics takes advantage of the LabVIEW program provides. It is mainly the fact that LabVIEW is a graphical programming language that uses icons instead of lines of text to create applications. In contrast to text-based programming languages, where instructions determine

program execution, LabVIEW uses dataflow programming, where the flow of data determines execution. In LabVIEW, you build a user interface with a set of tools and objects. The user interface is known as the front panel. You then add code using graphical representations of functions to control the front panel objects. The block diagram contains this code. In some ways, the block diagram resembles a flowchart.

2. Serial communication between LabVIEW and simple meter

Serial communication requires that you specify four parameters: the *baud rate* of the transmission, the number of *data bits* encoding a character, the sense of the optional *parity bit*, and the number of *stop bits*. Each transmitted character is packaged in a character frame that consists of a single start bit followed by the data bits the optional parity bit, and the stop bit or bits. A typical character frame encoding the letter “m” is shown here Lab (2006).

Fig. 1. Typical character frame encoding the letter “m”Lab (2006).

Baud rate is a measure of how fast data is moving between instruments that use serial communication. RS-232 uses only two voltage states, called MARK and SPACE. In such a two-state coding scheme, the baud rate is identical to the maximum number of bits of information, including “control” bits, that are transmitted per second.

MARK is a negative voltage and SPACE is positive; the previous figure shows how the idealized signal looks on an oscilloscope. The truth table for RS-232 is:

Signal > +3 V = 0

Signal < -3 V = 1

The output signal level usually swings between +12 V and -12 V. The “area dead” between +3 V and -3 V is designed to absorb line noise.

A start bit signals the beginning of each character frame. It is a transition from negative (MARK) to positive (SPACE) voltage; its duration in seconds is the reciprocal of the baud rate. If the instrument is transmitting at 9600 baud, the duration of the start bit and each subsequent bit will be about 0.104 ms. The entire character frame of eleven bits would be transmitted in about 1.146 ms.

“Data bits” are transmitted “upside down and backwards”. That is, inverted logic is used and the order of transmission is from least significant bit (LSB) to most significant bit (MSB). To interpret the data bits in a character frame, you must read from right to left, and read 1 for negative voltage and 0 for positive voltage. For the figure above, this yields 1101101 (binary) or 6D (hex). An ASCII conversion table shows that this is the letter “m”.

An optional *parity bit* follows the data bits in the character frame. The parity bit, if present, also follows inverted logic (1 for negative voltage and 0 for positive voltage.) This bit is included as a simple means of error checking. You specify ahead of time whether the parity of the transmission is to be even or odd. If the parity is chosen to be odd, the transmitter will then

set the parity bit in such a way as to make an odd number of 1's among the data bits and the parity bit. The transmission in the figure above uses odd parity. There are five 1's among the data bits, already an odd number, so the parity bit is set to 0.

The last part of a character frame consists of 1, 1.5, or 2 *stop bits*. These bits are always represented by a negative voltage. If no further characters are transmitted, the line stays in the negative (MARK) condition. The transmission of the next character frame, if any, is heralded by a start bit of positive (SPACE) voltage.

2.1 Serial port (RS-232)

The RS-232 is a standard developed by the Electronic Industries Association (EIA) and other interested parties, specifying the serial interface between Data Terminal Equipment (DTE) and Data Communications Equipment (DCE). The RS-232 standard includes electrical signal characteristics (voltage levels), interface mechanical characteristics (connectors), functional description of interchange circuits (the function of each electrical signal), and some recipes for common kinds of terminal to modem connections. The most frequently encountered revision of this standard is called RS-232C. Parts of this standard have been "adopted" (with various degrees of fidelity) for use in serial communications between computers and printers, modems, and other equipment. The serial ports on standard IBM compatible personal computers follow RS-232 Lab (2003).

Devices that use serial cables for their communication are split into two categories. These are DCE (Data Communications Equipment) and DTE (Data Terminal Equipment.) DCE are devices such as your modem, TA adapter, plotter, etc., while DTE is your computer or terminal. RS-232 serial ports come in two "sizes", the D-Type 25-pin connector and the D-Type 9-pin connector. Both of these connectors are male on the back of the PC; thus, you will require a female connector on your device. Below is a table of pin connections for the 9-pin and 25-pin D-Type connectors Lab (2006).

(a) DB-9 connector and pinouts.

(b) DB-25 connector and pinouts.

Fig. 2. Serial connectors

The 9 pin's connector is occasionally found on smaller RS-232 lab equipment. It is compact, yet has enough pins for the "core" set of serial pins (with one pin extra). Important: The DB-9 pin numbers for transmit and receive (3 and 2) are opposite of those on the DB-25 connector (2 and 3). Be careful of this difference when you are determining if a device is DTE or DCE. The 29 pin's connector is "standard" RS-232 connector, with enough pins to cover all the signals

Function	signal	9 pin	25 pin	DTE	DCE
Data	TxD	3	2	Output	Input
	RxD	2	3	Input	Output
Handshake	RTS	7	4	Output	Input
	CTS	8	5	Input	Output
	DSR	6	6	Input	Output
	DCD	1	8	Input	Output
	DRT	4	20	Output	Input
Common	Com	5	7	-	-
Other	RI	9	22	Input	Output

Table 1. Lines of serial interface Lab (2006).

specified in the standard. The table shows only the “core” set of pins that are used for most RS-232 interfaces Lab (2003).

2.2 Control of serial port in LabVIEW

The LabVIEW **Instrument I/O Serial** subpalette contains functions and VIs used for serial port communication. You should notice that some of the functions in this subpalette are the VISA functions you used previously for GPIB communication. The VISA Write and VISA Read functions will work with any type of instrument communication and are the same whether you are doing GPIB or serial communication. However, because serial communication requires you to configure extra parameters, you must start the serial port communication with the VISA Configure Serial Port VI.

Fig. 3. VISA configure serial port Lab (2006).

The VISA Configure Serial Port VI initializes the port identified by **VISA resource name** to the specified settings. **Timeout** sets the timeout value for the serial communication. **Baud rate**, **data bits**, **parity**, and **flow control** specify those specific serial port parameters. The **error in** and **error out** clusters maintain the error conditions for this VI Lab (2003).

The example above shows how to send the identification query command *IDN? to the instrument connected to the COM2 serial port. The VISA Configure Serial Port VI opens communication with COM2 and sets it to 9600 baud, 8 data bits, odd parity, one stop bit, and XON/XOFF software handshaking. Then the VISAWrite function sends the command. The VISA Read function reads back up to 200 bytes into the read buffer, and the error condition is checked by the Simple Error Handler VI.

Fig. 4. The example of configuration. Lab (2006).

2.3 Specific application of serial communication in LabVIEW

The aim of the study was to create a measurement programme within a Lab VIEW programming environment for a laboratory task of “efficiency measurement of photovoltaic conversion”. The use of the programme has brought the following benefits:

- digitalization of measurement,
- automation of measurement,
- archiving of measured values,
- easy further processability of values,
- immediate graphic processing,
- speed of measurement,
- and others.

2.3.1 Analysis of laboratory task

The programme should read the values of current and voltage on a photovoltaic cell. This is done with two Metex multimeters, one connected as a Voltmeter, the other as a Amperemeter. The values measured by the devices should be displayed on the computer. A good solution is to connect the measuring devices to the computer by means of a serial port. As there was not a computer available equipped with two RS-232 serial ports, it was necessary to use a Delock converter from a RS-232 interface to a USB for one device. Although the communication via the RS-232 interface compared to the more advanced USB is rather slow, it is wholly sufficient for this task.

The value from the measurement devices is displayed on the computer screen at once, without delay, and that is why the programme meets another requirement for the measurement, namely the recency of the values displayed. Data should further be processed in the structure of tables and simple charts. This requirement has been met by using appropriate blocks in the programme which as a result saves data in three columns of a table and then plots the current dependency charts: $P = f(I)$, $U = f(I)$ and $P; U = f(I)$.

Another requirement for the programme function was to provide a option of saving or exporting the measured data into a different programme. The measured data programme. This is enabled by one block that asks the user to enter the file name and path to the file. The data is saved in the way as shown in the table. The LabVIEW programme creates files with a LVM suffix (LabVIEW Measurement), that can be opened in any text editor, e.g. NotePad, MS Word.

2.3.2 Programme analysis

As mentioned above, programmes are created within Lab VIEW through merging blocks. In the programme for measurement only some of the great amount of those prepared in advance were used. Serial communication surely belongs to the essential parts of the programme. There is more than one way within Lab VIEW to effecting serial communication within the measuring device and the computer. Within the Functions -> Instrument I/O -> Serial palette we will find nodes for communication with a serial port. These are now referred to as “old-fashioned”. Their use is no longer recommended at present. They are included in the palette to provide compatibility with the earlier applications. The author had difficulty operating the control lines of the serial port using these functions. The other possibility of operating the port is to use e.g. library .dll which exports the relevant functions. The author, though, has not used the mentioned procedure either.

The National Instruments company has developed a higher level comprehensive interface for the control of peripherals, named NI-VISA. This is basically a set of what is called “API functions”, that can be generally called from almost any programmable language and that can use a wide range of differing instruments, i.e. by far not just the serial port. There is also a VISA interface for LabVIEW, i.e. for the language G. National Instruments recommends using the latter interface for the new applications.

2.3.3 Communication with the measuring device

The serial port is a system means. In order to be able to use it, we need to designate it for use in our application. This phase of the port use is called “opening” of the port and in the block diagram it will be realized by the Functions -> Instrument I/O -> VISA -> VISA Advanced -> Visa Open node.

VISA Open (see Fig. 5) - opens sessions for equipment specification and returns a session identifier which can be used to call some other operations of this equipment.

Fig. 5. VISA Open with VISA Resource Name constant.

Relevant input parameters:

A VISA resource name - indicates category and source of the open VISA sessions. The session does not have to be the correct one. Only the category of sessions and the resource name are relevant.

The relevant output parameters:

A VISA resource name - is the source where a session and its category are opened. The category has to match the input VISA resource name.

A VISA Resource Name constant will be created through right-mouse clicking on the node and selecting the Create Constant option on the menu that will appear. Then we will define the value of the constant in the pulldown menu. We will be able to choose from several options, depending on what interfaces are installed in the system at that moment. ASRL1::INSTR is the code for the first serial port, and similarly ASRL2::INSTR codes the other serial port.

Next the communication parameters will need to be set, i.e. the Baud Rate (speed of communication) to 1200, the number of Data Bits to 7, the number of Stop Bits to 2.0, and the level of RTS State control line to Unasserted. We will set the parameters for the open port using the Functions - Instrument I/O -> VISA - VISA Advanced -> Property Node.

The Property Node (see Fig. 6) - This node has an adjustable number of input and output terminals. In the individual Property Node arrays the required property can be set up. Each array is either designed for reading or for entry. We will set this property in the menu by the right - mouse clicking ("Change to Read", resp. "Change to Write" items). To summarize, we will choose the needed properties, set them for writing and assign values to them using the constants.

Fig. 6. Property Node with the set transmission parameters.

After that we will use the WHILE cycle and repeatedly execute entering of the CD command to the port and reading of the string (fourteen symbols for the Metex multimeters) from the port. The Functions -> Instrument I/O -> VISA -> VISA Read node is used for reading and the Functions -> Instrument I/O -> VISA -> VISA Write node for writing.

VISA Read (see Fig. 7) - the function is reading the prescribed number of bytes from the designated device or interface.

Fig. 7. VISA Read.

Relevant input parameters:

A VISA resource name - determines which interface will be opened. This parameter has the defined values of the ports named COM1, COM2 and LPT1.

Byte count - is the amount of bytes to be loaded to the buffer - 14 in our case.

Dup VISA resource name is a copy of the VISA resource name and is transmitted to the next VISA function.

Relevant input parameters:

Read buffer - transmits the symbols read from the entered interface.

VISA Write (see Fig. 8). Enters data from the "write buffer" to the device or interface specified in the VISA resource name.

Fig. 8. VISA Write.

Relevant input parameters:

write buffer - contains data to be entered in the device.

Dup VISA resource name is a copy of the VISA resource name and is transmitted to the next VISA function.

The WHILE cycle ends once the user has clicked the Stop button which is a representation of the terminal of the same name in the block diagram. The node D is responded by the Controls -> String & Path -> String Indicator (see Fig. 9) on the front panel, where the received string is displayed.

Fig. 9. String Indicator.

When using the port is at an end, it should be closed for other applications to be able to gain access to it if they request to. The closing of the port is done through the Functions -> Instrument I/O -> VISA -> VISA Advanced -> Visa Close node.

VISA Close (see Fig. 10) closes the specified session of the device or the resulting object. VISA Close accepts all the available categories.

Fig. 10. Visa Close.

The general view of the serial port operation:

We have two measuring devices connected to the computer over two serial ports. Within the programme the serial port operation has to be executed twice. The simplest serial communication will be activated by just connecting of all the aforementioned blocks in a row. The block connection of the serial port operation for the COM2 port is shown below (see Fig. 11):

Fig. 11. Operation of Serial Port.

2.3.4 Data processing, calculations

We have activated the serial port in the Lab VIEW environment and the measuring devices have already been sending data to the programme in regular intervals. The data is, however, of the String data type which cannot be used for calculations and so the data type needs to be changed. Moreover, the string sent by the multimeter also contains an abbreviation relating to the measured quantity and a unit of the given quantity, blocks for the transmission of the data type would not be able to convert the values and it is therefore necessary to remove all symbols except for digits from the string.

Match Pattern (see Fig. 12) is an appropriate block for removing of selected symbols from the *string*. The input parameters of the block include a string, where a string is loaded for filtering, and a *regular expression* that contains symbols that will be retained in the string. In our case we wish to leave all the digits and a sign in place, the regular expression therefore contains a

string of the necessary symbols. The resulting output is a *match substring* which only contains digits and a sign.

Fig. 12. Match Pattern.

In order to convert a string to a “double” data type used by all the computing blocks, we will use the Fract!Exp String To Number block (see Fig. 13). We will send a string to the *string* input and on the *number* output we will obtain a value of the measured quantity in the numerical form that we can already use for calculations.

Fig. 13. Fract/Exp String to Number.

From the serial ports' outputs we have obtained the Voltage U and Current I values, flowing through the FV (photovoltaic) cell. We will calculate the Performance P value by using the Multiply block (see Fig. 14) which can perform multiplication between the numbers brought to the input x and y , and their product then appears on the output.

Fig. 14. Blok Multiply.

We will set the value of an impact performance P_{dopad} , to be calculated based on pyranometer sensitivity and pyranometer output voltage, using the Numeric Control (see Fig. 15) block. This we will find within Controls - Numeric Controls - Numeric Control on the Front Panel. Similarly, we will also set the size of area S .

Before calculating efficiency a maximum performance P_{max} of a FV cell yet needs to be determined. We will obtain the maximum value by using a Max&Min block (see Fig. 16). The

Fig. 15. Numeric Control on the Front Panel and Block Diagram.

Fig. 16. Max&Min.

Max&Min block downloads the values transferred to its inputs, which results in the value on the Max output.

Using the Multiply block (see Fig. 14) we will determine the area capacity rating P_{plon} as the product of a maximum performance P_{max} and a cell area S .

The values P_{dopad} and P_{plon} are known to us, and using the Divide block we can determine the efficiency of a FV cell. The Divide block (see Fig. 17) executes the quotient of two values transferred to its inputs, and the output shows the quotient as the result. The quotient and the product could also be determined using the Formula block, but the Multiply and Divide blocks are sufficient for calculations this simple.

Fig. 17. Blok Divide.

The immediate value of efficiency η , along with the values of Current I , Voltage U , Performance P , Cell Area S and Impact Performance P_{dopad} can be displayed on the front panel (see Fig. 18).

The front panel further consists of a table to which we enter the measured values. The block for the table is placed in Controls -> Text indicators -> Table (see Fig. 19). In order to be able to write at the same time the values of current, voltage and performance to the table, we need to use the Merge Signals (see Fig. 19) block which will merge the signals into a single one which we will bring to the single input of the table. We can set the number of columns and rows in the table, in our case 3 columns and 20 rows. The options further provide for the display of headings for columns and rows and of a scroll-bar.

Fig. 18. Current values on the Front panel.

Fig. 19. Table of values with Merge Signals block and tabs for entry and deletion.

On the front panel of the table we can see the values of current, voltage and performance in the columns (see Fig. 20). We will enter values into the table by bringing Log 1 to the Enable input, and perform the table deletion by bringing Log 1 to the Reset input. These operations are executed by means of two tabs on the Front panel, to be found in the Controls -> Buttons & Switches -> Text Button (see Fig. 20).

We can plot the measured values in charts; there were three dependence charts needed for this task, specifically VA - characteristic $U = f(I)$, Performance characteristic $P = f(I)$ and Performance as well as VA characteristics in a single chart $U, P = f(I)$. To create the charts we chose three Build XY Graph blocks (see Fig. 21), to be found in the Controls -> Graph Indicators -> XY Graph. Each block of the chart has two inputs for the values x and y on the axis and further inputs for enabling Entry and Deletion, that we will connect similarly as for the block of the table, to the tabs ENTER and DELETE.

Tabulka hodnot

U [V]	I [mA]	P [mW]
0,439000	8,600000	3,775400
0,426000	20,100000	8,562600
0,404000	36,300000	14,665200
0,399000	38,900000	15,521100
0,392000	43,000000	16,856000
0,384000	47,200000	18,124800
0,373000	52,400000	19,545200
0,356000	58,200000	20,719200
0,333000	64,900000	21,611700
0,317000	68,500000	21,714500
0,306000	70,800000	21,664800
0,297000	72,100000	21,413700
0,277000	75,300000	20,858100
0,255000	77,000000	19,635000
0,245000	78,500000	19,232500
0,211000	80,100000	16,901100
0,192000	82,500000	15,840000

Zapsat

Vymazat

Fig. 20. Table with measured values, tabs for entering values and deletion on the right.

Fig. 21. Build XY Graph block with entry and deletion tabs.

The window of the chart on the front panel is displayed as shown (see Fig. 22). In order to be able to display multiple courses on the chart $U, P = f(I)$, we need to again use the Merge Signals block and select two axes and two courses in the chart setup. The chart options include the setting of the cursor and further the setting of colours of the courses/curves and the cursor. The chart can connect points (values) in several ways, though any connecting line of a trend (linear, logarithmic) cannot unfortunately be used here, as provided e.g. within the MS Excel programme.

After the measurement of data the next step involves its saving. The Write LabVIEW Measurement File block (see Fig. 23) was used to save the data in a file. This block is located in the Function -> Output -> Write LabVIEW Measurement File in the Block diagram. I have connected the writing to the table, to the charts and the file to a single tab, since these operations should be started at the same point in time. The programme should save data

Fig. 22. Grafy na předním panelu.

to a single file, give the user a possibility of choosing a path to the file and the file name, and that is why I have chosen to use 'Action -> Save to one file -> Ask user to choose file -> Ask only one' in the block setting. During the first measurement a dialog box appears to save a file (see Fig. 23). Within the block setting you can further select display of a heading for a file and time of measurement. To delete the content of a file we use the DELETE tab, which is common for the table and the chart. The LabVIEW programme creates files with a LVM suffix (LabVIEW Measurement), that can be opened in any text editor.

Fig. 23. Dialog box for writing to a file.

Overall view of the front panel and the block diagram is shown below.

Page 1

pokus3.vi
 D:\Texty\texty-win\Clanky\2010\In-TECH\Zdrojaky\FV_kanotr\pokus3.vi
 Last modified on 31.5.2007 at 11:46
 Printed on 23.9.2010 at 8:06

Front Panel

Fig. 24. General view of the front panel of a virtual device.

Page 1

pokus3.vi
 D:\Texty\texty-win\Clanky\2010\In-TECH\Zdrojaky\FV_kanotr\pokus3.vi
 Last modified on 31.5.2007 at 11:46
 Printed on 23.9.2010 at 8:08

Block Diagram

Fig. 25. Block diagram of a virtual device.

This is an example of a functioning virtual device which communicates through a serial line with a real measuring device, visualizes the measured values and further processes them. This is one of the basic uses of Lab VIEW for the construction of a practical application. A device constructed in this way, or a source code, can be further compiled to form a separate application independent of the LabVIEW programming environment. The following picture is just a sample design of Metex, a simple virtual device communicating through a serial port.

Fig. 26. Block diagram of a Metex 4560 cr. virtual device.

VISA configurator module is used again for communication via a serial port, which is, however, part of a “sub vi” application. The other blocks are just the means to process, control and display data. A detailed analysis of this application, however, is not the focus of this paper.

3. Flickermeter in the LabVIEW environment

3.1 Introduction

One of the major problems in regard to the quality of electric power involves quick changes in the voltage, i.e. its fluctuation, causing flickering of the light flow from the light sources. The light flow flickering may subsequently create an undesired perception of flickering which adversely impacts visual performance. This is a subjective visual impression initiated by a light stimuli, whose brightness or spectral distribution changes in time. The flickering of light sources may cause mental or physical fatigue or major physiological discomfort in people.

Voltage fluctuation may be caused by interference during the generation, transmission, distribution and consumption of electricity. The voltage fluctuation in the grid may be due especially to the fluctuation of active and reactive power of some appliances.

In order to measure the degree of perception of flickering was developed a UIE/IEC flickermeter which based on the measurement of voltage fluctuations evaluates the corresponding level of perception of flickering. The flickermeter models the transmission

The flickermeter described in the standard is an analogue measuring device. The present-day digital measuring devices use a modified version of the standard flickermeter. S-transmission functions of analogue filters, used for describing the device in the standard, have been transformed into a digital form using bilinear Z-transformation.

UIE / IEC flickering-meter model was constructed as a virtual device within LabView. It is made up of the following basic blocks:

- voltage measurement,
- input signal normalization,
- filtration and weighting,
- statistical analysis.

Fig. 28. Block diagram of a virtual device.

3.3 Voltage measurement

The voltage measurement uses the NI-PXI 4472 measurement card which is able to measure voltage in the range $\pm 10V$. To measure the mains voltage it is necessary a ballast that will modify the magnitude of voltage; Slezinger et al. (2009).

The card's inputs are galvanically separated and properly isolated so it is possible to bring the mains voltage potential to the input terminals. This allow for using the voltage modification with a resistance voltage divider proven to provide the highest accuracy and linearity in the broad frequency range.

Voltage measurement for the flickermeter in the LabView block diagram is taken by means of express VI DAQ Assistant from the Input palette. The assistant automatically finds the measurement cards connected to the computer and mediates their configuration by the user. The measurement card is then represented by a single block in the block diagram, whose output is a data flow with an excessive quantity.

The measurement card connection is detailed in the LabView Help. For the purpose of flickermeter operation it was necessary to set an "Acquire Continuously" option in the Task Timing tab. A sampling frequency is set for the required 3200 S/s. The sampling frequency selection has to be made right at the start of flickermeter designing as all the filters used are specified using constants that depend on the sampling frequency.

3.4 Simulation of the light source - eye - brain chain

The block of signal normalization includes the measurement of input voltage effective value and calculates a maximum value. The input signal is normalized to a unit-amplitude signal.

The simulation of the light source-eye-brain chain is performed as a cascade of filters. The filters used in the flickermeter are designed as a cascade of digital filters. There is IIR Filter used from the Signal Processing -> Filters -> Advanced IIR Filtering menu. The transmission function of this filter is specified using Forward Coefficients and Reverse Coefficients in the form of array of coefficients. The array is set up using the Build Array block.

A "TRUE" signal to the init/cont input is also transferred to the input of the IIR Filter block, for the signal filtering to go on continually for the whole period of the programme run, Hanzelka & Bien (2000).

The complete solution of one of the filters is shown in (see Fig. 29).

Fig. 29. Block diagram of the constructed filter.

3.4.1 Problems relating to filter stability

The practical implementation of the digital filters within LabView has indicated an increased likelihood of filters becoming unstable if the transmission function of the filter is higher than the second order of magnitude. It is therefore useful to transform transmission functions of the higher orders to a product of transmission functions of the lower orders of magnitude.

3.4.2 The programme run

The complete flickermeter is created as a programme loop named WhileLoop with 1 s programme delay using the TimeDelay block. Within each step of the loop a block of 3200 samples is read from the measurement card and data is transformed from the DynamicData type to the Array data type which is processed by individual filters. The filters' output is a single value for each block, denoted as P_f , of immediate degree of perception of flickering. P_f values are saved in an accumulator designed as Express VI Collector, during each individual passage of the loop. It accumulates values for a period of time required for calculation of the PST value (typically 10 minutes, though the programme allows for different times of accumulation. too).

3.4.3 Statistical signal analysis

An array of values is again placed on the accumulator output, administered in a statistical classifier whose block diagram is represented here below. The classifier function is consistent with the description within the standard. The number of classes required by the standard is minimally 64, the constructed algorithm uses 10000 of them.

Fig. 30. Statistical evaluation of data.

3.4.4 The flickermeter control

START and STOP pressbuttons and Pf and Pst indicators are placed on the front panel. The input voltage curve, Pf current value and Pst measured results are displayed on the graphic displays.

The text indicator displays information about the current status of the meter. Three possible states may take place:

- starting the meter with counting-down the time-period for which it has to wait for the initial transient performance to pass,
- “Wait” status when the flickermeter stands by and waits for the START button to be pressed. This status takes place after the set time for the meter’s starting lapsed, or after the STOP button is pressed,
- status of data accumulation for subsequent calculation of Pst value while counting down the remaining time of accumulation. This state stays active after pressing the START button until the STOP button is pressed, or the time of accumulation has lapsed, and the Pst calculation is made. The Pst value will always be overridden with the latest value on the display after the time of accumulation lapsed.

The flickermeter is programmed so as to enable using of the measured values also as a data source for further VI running on the same computer. The flickermeter is part of a comprehensive measurement system that can remote start the flickermeter and read the measured results using global variables.

Fig. 31. Block diagram of the timer.

The general view of the user interface of the flickermeter is as shown on the following page.

Fig. 32. Block diagram of the timer.

The flickermeter has been verified by the tests required by the standard. Its response to the test signals is fully compliant with its requirements and it now used within the development of the new types of the meter.

The described meter is now used as a reference device in the research of new constructional solutions of the flickermeter and in the development of a generator of different kinds of disturbance voltage fluctuations. It can be operated on any kind of computer with a measurement card connected. It can be built into comprehensive systems focusing on the area of electric power quality, based on virtual instrumentation.

4. Communication and control of Viasala WXT520 meteostation in LabVIEW environment

Finally the communication of LabView and a Viasala WXT520 meteostation is described. The meteostation is part of an extensive measurement system which on a long-term basis monitors electrical and physical quantities of the system of solar panels. A picture of a meteostation is shown in 33.

The meteostation delivery also included configuration software which, though, can only be used for meteostation configuration and testing of meteostation. With regard to the needs of the considered system of long-term online monitoring of data from solar panels the delivered software was inadequate as it did not enable digital access to the measured data. It was therefore needed to define and create our own application that would allow for integrating the WXT520 meteostation into the measurement system.

The core of the problem thus was to create an application within LabVIEW, that would be able to communicate with the meteostation using the RS-232 serial interface and would transfer the measured data further on, making it available for further processes within the measurement system that has also been programmed in the LabVIEW development programme environment.

The unique feature of this application is that it enables connection of a device to LabVIEW which does not have a primary support within LabVIEW in the form of measurement card, proving universality of the LabVIEW software environment. The only supported common communication standard is the serial port RS-232 interface.

The meteostation includes numerous detectors and sensors (see Fig. 33) measuring temperature, pressure, speed of wind, rainfall etc.

where:

- 1 wind transducers (3 pcs)
- 2 precipitation sensors
- 3 pressure sensor inside PTU module
- 4 Humidity and temperature sensors inside the PTU module

The basis for the communication configuration is the table of wiring of screw terminal pins on the meteostation terminal block, based on which the VISA block was set in the block diagram of the measurement application.

Fig. 33. Meteostanice WXT520 Viasala VIA (2008).

Fig. 34. Cut away view VIA (2008).

Screw Terminal Pin	RS-232	SDI-12	RS-485	RS-422
1 RX-	-	-	Data-	Data in (RX-)
2 RX+	-	-	Data+	Data in (RX+)
3 TX-	Data out (TxD)	Data in/out (Tx)	Data-	Data out (TX-)
4 TX+	-	-	Data+	Data out (TX+)
5 RxD	Data in (RxD)	Data in/out (Rx)	-	-
6 SGND	GND for data	GND for data	-	-
17 HTG-	GND for Vh+	GND for Vh+	GND for Vh+	GND for Vh+
18 HTG+	Vh+ (heating)	Vh+ (heating)	Vh+ (heating)	Vh+ (heating)
19 VIN-	GND for Vin+	GND for Vin+	GND for Vin+	GND for Vin+
20 VIN+	Vin+ (operating)	Vin+ (operating)	Vin+ (operating)	Vin+ (operating)

Fig. 35. Connection of pins for different VIA communication interfaces VIA (2008).

As can be seen from the block diagram of the application (see Fig. 37), it was necessary to configure a number of further variables to ensure the correct function, such as the following:

- channel selection,
- measurement interval,
- input port,
- communication protocol.

Subsequently the complete programme was created, whose user interface is shown below (see Fig. 36). The logic of the programme is roughly outlined by the following:

- setting of inputs
 - on/off measurements
 - definition of channels, ports etc.
- definition and configuration of individual sensors,
- repeated reading of data from the meteo-station in set intervals,
- display and entry of values.

5. Conclusion

The paper describes a few examples of using LabVIEW for the measurement of electric and non-electric quantities in electro-energy applications. The most frequently involved is evidently the LabVIEW communication with a measuring device using a serial port, but there are also applications putting it to good use, such as the described flickermeter which is constructed directly on the NIPXI platform, distributed by the National Instruments company. Through virtual instrumentation and launching of LabVIEW the simplification and greater effectiveness of the design have been achieved, and the same has in turn been applied in the construction of new laboratory tasks and research measurements. The described examples

Fig. 36. User interface.

indicate possible use of LabVIEW as a concentrator of individual measuring devices and peripherals, which can be customized based on user requirements. Its operation is based on a single measurement interface which combines all the varied instruments used, and visualizes, archives and processes the measured data and if necessary also transmits it to another part of the measurement system. An integral part of the programme is the possibility of creating an effective user interface, from which the system can be administered as a whole. The final part of the paper also deals with a possibility of connecting LabVIEW to and its communication with a non-standard instrument which is not primarily designed for the communication with LabVIEW. As it has turned out, provided that standards of the communication interface are complied with, in our case the RS 232 protocol, the connection can be basically made without difficulty. This makes LabVIEW a universal software in the sphere of virtual instrumentation.

6. Acknowledgment

This paper includes the results of research funded by the Ministry of Education, Youth and Sports of the Czech Republic as part of project No. MSM 0021630516 and project No. CZ.1.05 /2.1.00/01.0014 Centre for Research and Utilization of Renewable Energy.

7. References

Hanzelka, Z. & Bien, A. (2000). Guide to quality of electrical supply for industrial installations, part 5, flicker and voltage fluctuations, *Technical report*, Power Quality - Working Group WG2, Tomsk.

Fig. 37. Block diagram of the communication and control programme .

- IEC (1997). 61000-4-15 - cem - part 4: Testing and measurement techniques - section 15: Flickermeter - functional and design specifications, International Electrotechnical Commission.
- Lab (2003). *LabVIEW User Manual*, edition - part number 320999e-01 edn.
- Lab (2006). *LabVIEW™ Basic I Introduction Course Manual*, edition edn.
- Slezingr, J., Drapela, J. & Tezky, J. (2009). Flickermeter implementation in labview, *Proceedings of the 15th International Scientific and Practical Conference of Students, Post- graduates and Young Scientists MODERN TECHNIQUE AND TECHNOLOGIES MTT 2009*, Tomsk, pp. 11–13.
- VIA (2008). *User's Guide*, edition - m210906en-a edn.

Modeling, Programming and Simulations Using LabVIEW™ Software

Edited by Dr Riccardo De Asmundis

ISBN 978-953-307-521-1

Hard cover, 306 pages

Publisher InTech

Published online 21, January, 2011

Published in print edition January, 2011

Born originally as a software for instrumentation control, LabVIEW became quickly a very powerful programming language, having some characteristics which made it unique: simplicity in creating very effective User Interfaces and the G programming mode. While the former allows for the design of very professional control panels and whole applications, complete with features for distributing and installing them, the latter represents an innovative way of programming: the graphical representation of the code. The surprising aspect is that such a way of conceiving algorithms is extremely similar to the SADT method (Structured Analysis and Design Technique) introduced by Douglas T. Ross and SofTech, Inc. (USA) in 1969 from an original idea by MIT, and extensively used by the US Air Force for their projects. LabVIEW enables programming by implementing directly the equivalent of an SADT "actigram". Apart from this academic aspect, LabVIEW can be used in a variety of forms, creating projects that can spread over an enormous field of applications: from control and monitoring software to data treatment and archiving; from modeling to instrument control; from real time programming to advanced analysis tools with very powerful mathematical algorithms ready to use; from full integration with native hardware (by National Instruments) to an easy implementation of drivers for third party hardware. In this book a collection of applications covering a wide range of possibilities is presented. We go from simple or distributed control software to modeling done in LabVIEW; from very specific applications to usage in the educational environment.

How to reference

In order to correctly reference this scholarly work, feel free to copy and paste the following:

Jan Machacek, Jan Slezinger and Jiri Drapela (2011). Electric and Nonelectric Quantities Measurement in Power Engineering Using LabView, Modeling, Programming and Simulations Using LabVIEW™ Software, Dr Riccardo De Asmundis (Ed.), ISBN: 978-953-307-521-1, InTech, Available from:
<http://www.intechopen.com/books/modeling-programming-and-simulations-using-labview-software/electric-and-nonelectric-quantities-measurement-in-power-engineering-using-labview>

INTech
open science | open minds

InTech Europe

University Campus STeP Ri
Slavka Krautzeka 83/A
51000 Rijeka, Croatia

InTech China

Unit 405, Office Block, Hotel Equatorial Shanghai
No.65, Yan An Road (West), Shanghai, 200040, China
中国上海市延安西路65号上海国际贵都大饭店办公楼405单元

Phone: +385 (51) 770 447
Fax: +385 (51) 686 166
www.intechopen.com

Phone: +86-21-62489820
Fax: +86-21-62489821

© 2011 The Author(s). Licensee IntechOpen. This chapter is distributed under the terms of the [Creative Commons Attribution-NonCommercial-ShareAlike-3.0 License](#), which permits use, distribution and reproduction for non-commercial purposes, provided the original is properly cited and derivative works building on this content are distributed under the same license.