In-Vitro Magnetoresistive Biosensors for Single Molecular Based Disease Diagnostics: Optimization of Sensor Geometry and Structure

Seongtae Bae

*Department of Electrical and Computer Engineering, Biomagnetics Laboratory (BML),
National University of Singapore, 117576, Singapore
Singapore*

1. Introduction

Detection of biological information using magnetoresistance (MR) sensors based on multilayered giant MR (GMR), or exchange biased GMR spin valves, operated by the magnetic field produced by magnetic nanoparticle sensor agents has been paid considerable attention in biomedical sensor technologies (Baselt et al., 1998; Tondra et al., 2000; Rife et al., 2003; Graham et al., 2004). The main reason for this interest is that MR based biosensors provide technical advantages such as high sensitivity, relatively fast and low-volume assay, and easy manipulation of the magnetic sensor agents under the externally applied magnetic field gradients (Graham et al., 2002; Lagae et al., 2002). These advantages lead to improved sensing performance, stimulating the development of functional *in-vitro* GMR based biosensors to obtain biological information and diagnose diseases more accurately in healthcare (Megen & Prins, 2005).

However, all the developed GMR biosensors so far were mostly focused on counting or indentifying multiple biomolecules such as a DNA counter and a bead array counter rather than single molecular detection (SMD) (Baselt et al., 1998; Miller et al., 2001; Graham et al., 2003; Li & Wang, 2003; Schepper et al., 2004; Wang et al., 2005; Shen et al., 2005). The main physical reason for this technical limitation is that the magnetic susceptibility of a single superparamagnetic nanoparticle sensor agent (SPNSA), which can be easily manipulated on the sensor surface without any serious agglomeration and easily retrieve magnetic stray field, is too small to generate a large enough field for achieving a reasonably high SNR (Signal-to-Noise Ratio). Although superparamagnetic microbeads have been attempted to obtain a larger stray field for higher SNR for SMD (Graham et al., 2002; Wirix-Speetjens et al., 2006), these agents were also revealed to have a technical drawback that they can not maintain one to one ratio between the microbead and the biomolecules due to a big mismatch in size. Hence, to develop more powerful *in-vitro* GMR biosensor system for SMD, ferrimagnetic nanoparticle (FN) sensor agents, which have a high remnant magnetization expecting to producing a sufficient stray field for a higher SNR, and a high chemical stability as well as a high biocompatibility with living cells or biological entities, are currently considered as a feasible sensor agent to label the biomolecules. However, the applications of FN to a GMR biosensor agent has been limited for the past few years by a

Source: Intelligent and Biosensors, Book edited by: Vernon S. Somers,et,
ISBN 978-953-7619-58-9, pp. 386, January 2010, INTECH, Croatia, downloaded from SCIYO.COM

technical challenge relevant to a possible particle agglomeration while they are introducing into the microchannel. Fortunately, as a new functional microfluidic channel with micromagnet, which can allow for effectively manipulating the FNs to flow into the channel one by one, has been recently developed (Lagae et al., 2002; Ramaden et al., 2006; Latham et al., 2007), the interests to apply a single FN sensor agent to an in-vitro GMR biosensor for SMD are dramatically increased in a molecular diagnostic biosensor system.

In this chapter, the physical characteristics of an in-vitro GMR biosensor with an immobilized single FNSA is mainly discussed to provide crucial information how to optimize its structure for SMD. In chapter 2, the sensing performance of an in-vitro GMR biosensor with an immobilized FNSA or SPNSA is numerically analyzed based on the "Stoner-Wolfarth model" to evaluate which GMR biosensor system is more suitable for SMD. In chapter 3, the optimization of sensor geometry considering the spatial magnetic field interaction between the FNSA and the FL as well as the physical correlation between the "effective sensing area" and SNR is introduced to successfully design an in-vitro GMR biosensor enabling to show maximized SNR for SMD. Finally, chapter 4 discusses the newly designed sensor structure of an in-vitro GMR biosensor with a specially designed magnetic shield layer (MSL) to explore its technical effectiveness for the diagnostic biosensor applications based on SMD.

2. An in-vitro GMR biosensor with an immobilized single FNSA or SPNSA for SMD

To confirm which sensor system is more promising for SMD, the sensing performance of an in-vitro GMR biosensor with an immobilized FNSA or SPNSA should be first clarified. Four physical sensing parameters: (1) the relative MR change (δR) relevant to the magnetization angle difference between the free layer (FL) (sensing layer) and the pinned layer in GMR spin-valves, (2) the interaction factor (IF) describing the magnetic spatial interaction between the FL and the nanoparticle sensor agent, (3) the distance between the single immobilized FNSA or SPNSA and the FL, and (4) the practically allowable sensor size considering the physical limit of current sensor fabrication technology, are mainly considered for the comparison of sensing performance. The numerical analysis is based on the "Stoner-Wolfarth model" describing that the magnetizations of the FL in the GMR biosensors are coherently rotated under a globally applied magnetic field with two-dimensional field components, which are homogeneous across the entire FL.

2.1 Physical model of spatial field distribution on an in-vitro GMR biosensor

The sensing mechanism of a GMR biosensor with an immobilized single FNSA or SPNSA is illustrated in Fig. 1-(a). The sensor agents are captured on the FL surface of GMR biosensor through a bio-recognition process and the resistance of the GMR sensor is accordingly changed by the stray field produced from the magnetic nanoparticles. As shown in Fig. 1, the critical geometry parameters considered for the numerical calculation include the spacing between the SA and the sensor surface (h), and the geometry of FL, e.g. length (L), width (W), and thickness (t). Figure 1-(b) describes the magnetization configuration of FL in the GMR biosensor and the magnetic spatial interaction between the SA and the FL. The magnetization of pinned layer (PL) is exchange biased to the negative transverse direction ($-y$) and the easy axis (EA) of the FL is along the longitudinal direction ($+x$) due to shape anisotropy and crystalline anisotropy. To minimize the geometrically-induced magnetostatic coupling

between the FL and the orthogonally coupled PL, the PL is considered as the “synthetic anti-ferromagnetic (AFM) layer”, e.g. FM/non-magnetic/FM structure. The magnetization of FNSA is aligned in the positive transverse direction (+y) and the SPNSA is magnetized toward the positive transverse direction (+y) by the excitation field (+y) for comparison with the FNSA. In order to establish a physical model for the stray field induced by the magnetic nanoparticles, the single magnetic nanoparticle with spherical shape is assumed to be uniformly magnetized and has an effective dipole magnetic moment (m). The field intensity, $H(H \cong B$ in cgs unit) at a distance r can be expressed by equation (1) (Besse et al., 2002).

$$\overrightarrow{H(B)} = \frac{3\hat{r}(\hat{r} \times m) - m}{r} \quad (1)$$

The spatial distributions of the stray field induced by the magnetic nanoparticle on the FL of GMR biosensor are numerically calculated by considering the both longitudinal and transverse magnetic field components of an ideal dipole, as illustrated in Fig. 1-(b).

$$H_x(B_x) = m \cdot \frac{2x^2 - y^2 - z^2}{(x^2 + y^2 + z^2)^{5/2}}$$

$$H_y(B_y) = m \cdot \frac{3xy}{(x^2 + y^2 + z^2)^{5/2}} \quad (2)$$

$$H_z(B_z) = m \cdot \frac{3xz}{(x^2 + y^2 + z^2)^{5/2}}$$

Fig. 1. A schematic diagram of an in-vitro GMR biosensor with a single nanoparticle sensor agent for SMD (a) the molecular recognition process on the GMR biosensor surface, (b) the magnetic spatial interaction between the magnetic nanoparticle agents and the FL of GMR biosensor. All the field components generated from the bias field, H_b , the excitation field, H_t and the magnetic nanoparticles are illustrated on the FL sensor surface, and (c) a schematic diagram of magnetization configuration of a single magnetic nanoparticle and its variation due to the magnetic dipole interaction caused by the x-component of FL magnetization to the FNSA or the SPNSA.

The total free energy in the FL is described by equation (3) based on the "Stoner- Wolfarth model" (Stoner & Wolfarth, 1948).

$$E = K_u \sin^2 \theta + \frac{1}{2}(N_z - N_x)M_s \sin^2 \theta - HM_s \cos(\psi - \theta) - H_t M_s \sin \theta - H_b M_s \cos \theta \quad (3)$$

Where H is the stray field generated by the magnetic nanoparticles, which can be divided into H_x and H_y , respectively. The demagnetizing factors can be determined by equation (4) (William, 2001).

$$N_z = 8tw/l\sqrt{l^2 + w^2}, \quad N_x = 8tw/w\sqrt{l^2 + w^2} \quad (4)$$

According to the "Stoner- Wolfarth model", the magnetization direction of the FL (FL is in a single domain state as the sensor size is in submicron range in this model) is determined by the minimization of the total free energy as shown in equation (3). Hence, the total free energy with respect to θ (Fig. 1-(c)) is minimized. Our method is to set up a discrete array of θ . The range of this array is from 0 to 2π and the step size is 0.00001. With such small step size, the range of θ can be considered as a continuous range. By substituting the array into equation (3), the minimized energy and the corresponding θ can be obtained using a simple algorithm. The next step is to use the following equation (5) to calculate the magnetoresistance (MR) ratio based on the magnetization (spin) configuration of GMR biosensor where the magnetization direction of PL is exchange biased to a fixed direction (-y).

$$\frac{\Delta R}{R} = \left(\frac{\Delta R}{R} \right)_0 \frac{1 - \cos(\theta_f - \theta_p)}{2} \quad (5)$$

Where θ_f and θ_p are the angles of the FL, and the PL with respect to EA, respectively; $(\Delta R/R)_0$ is the MR ratio of the GMR biosensor when the FL and PL is antiparallel with each other, which is also the maximum MR ratio for a GMR biosensor. To evaluate the sensing performance, the relative MR change, δR , is needed to be defined by equation (6).

$$\delta R = \left| \frac{(\Delta R/R)_{without} - (\Delta R/R)_{with}}{(\Delta R/R)_0} \right| \times 100(\%) = \frac{1}{2} \left| \sin \theta_{f,without} - \sin \theta_{f,with} \right| \times 100(\%) \quad (6)$$

Where $(\Delta R/R)_{with}$ indicates the MR ratio of GMR biosensors due to the nanoparticle sensor agents immobilized on the surface of FL, $(\Delta R/R)_{without}$ indicates the MR ratio without nanoparticle, especially SPNSA, on the surface of FL. Moreover, in equation (6), the $\theta_{f,without}$ is the angle between FL magnetization and EA when no sensor agent is on the sensor surface and the $\theta_{f,with}$ is the angle between FL magnetization and EA when the sensor agent is on the sensor surface. For FNSA, the $\theta_{f,without}$ is zero, thus δR can be written by $\delta R = 0.5 \sin \theta_{f,with}$. For SPNSA, the $\theta_{f,without}$ is not zero due to the excitation field (see Fig. 1-(b)) and thus the δR can be rewritten by $0.5 \cos \theta_f \cdot d\theta$, where, $d\theta$ is the angle difference before and after the SPNSA captured on the sensor surface ($d\theta = \theta_{f,without} - \theta_{f,with}$). The fringe field from the FL of GMR biosensors is also considered to affect the magnetic properties of nanoparticle sensor agent in both magnetization direction and magnetic moment due to the magnetic dipole interaction between the FL and the nanoparticles (see Fig. 1-(b)).

Accordingly, the effect of fringe field from the FL on the δR is included in this model to precisely interpret the sensing performance. To find out how many percentages the FL fringe field would influence on the magnetic moment of the magnetic nanoparticles, the interaction factor (IF) defined as $\alpha = (1 - m' \cos \beta / m) \cdot 100(\%)$ is employed, where m' is the magnetic moment considering the effect of the FL fringe field, m is the original magnetic moment and β is the angle difference between the original magnetization of sensor agent and the rotated magnetization due to the fringe field from FL as shown in Fig. 1-(c). In this model, the FNSA is considered as a CoFe_2O_4 nanoparticle. It has a remnant magnetization of 22 emu/g and a diameter of 26 nm (see Fig. 2-(a) and (b)). The magnetic moment of the CoFe_2O_4 nanoparticle is calculated using $\hat{m} = 4\pi M_r V$, where M_r is the remnant magnetization and V is the volume of the CoFe_2O_4 nanoparticle. Since the CoFe_2O_4 nanoparticle has a large remnant magnetization, the excitation field (H_t) is not required, thus the equation (3) can be simplified by equation (7),

$$E = K_u \sin^2 \theta_f + \frac{1}{2}(N_z - N_x)M_f \sin^2 \theta_f - H_x M_f \cos \theta_f - H_y M_f \sin \theta_f - H_b M_f \cos \theta_f \quad (7)$$

Due to the fringe field from the FL of GMR biosensor (ΔH), the induced magnetic moment of the CoFe_2O_4 nanoparticle becomes $\Delta m = \chi_f \cdot \Delta H$, where χ_f is considered as constant because the fringe field is relatively small. In addition, by considering the single domain state of CoFe_2O_4 nanoparticle agent, the magnetization direction of the CoFe_2O_4 nanoparticle can be calculated using the "Stoner- Wolfarth model" as below:

$$E = \frac{1}{2}(H_u + H_d)m' \sin^2 \beta - \Delta H_x m' \cos \beta - \Delta H_y m' \sin \beta - H_b m' \cos \beta \quad (8)$$

$$\Delta H_x = \frac{M_f l w t \cos \theta_f}{(h + t/2 + D/2)^3}, \Delta H_y = \frac{M_f l w t \sin \theta_f}{(h + t/2 + D/2)^3}$$

where $m' = m + \Delta m$, and ΔH_x , and ΔH_y are the longitudinal, and transverse component of the FL fringe field, respectively. As the saturation magnetization of the CoFe_2O_4 nanoparticle sensor agent is almost the same as the bulk CoFe_2O_4 ferrite, the anisotropy constant, K_1 , of the sensor agent is assumed as bulk value of CoFe_2O_4 ferrite, which is a 2×10^6 erg/cm³ and thus $H_u = 2K_1/m'$. The demagnetizing factor is considered as $4\pi/3$ thus, $H_d = m' \cdot 4\pi/3$.

The SPN used in this model is also a single CoFe_2O_4 nanoparticle but the diameter is a 7 nm as shown in Fig. 2-(c) and (d). As can be seen in the inset of Fig. 2-(c), the magnetic susceptibility, χ , of the superparamagnetic CoFe_2O_4 nanoparticle is almost constant at a 0.041 independent of applied magnetic field. Using the experimentally obtained χ value, the magnetic moment of the superparamagnetic CoFe_2O_4 nanoparticle under the H_b and the H_t is determined at $\hat{m} = 4\pi^3 \chi H_b x/3 + 4\pi^3 \chi H_t z/3$ and the total free energy can be correspondingly re-written by equation (9).

$$E = K_u \sin^2 \theta_f + \frac{1}{2}(N_z - N_x)M_f \sin^2 \theta_f - H_x M_f \cos \theta_f - H_y M_f \sin \theta_f - H_b M_f \cos \theta_f \quad (9)$$

$$- H_t M_f \sin \theta_f$$

As the χ value is constant and there is no magnetic anisotropy energy under no excitation field, the IF for the superparamagnetic CoFe_2O_4 nanoparticle sensor agent can be simplified as $\alpha_y = \Delta H_y / H_t$, where ΔH_y is the y component of the FL fringe field from GMR biosensor.

$$\alpha = \frac{M_s l w t \sin \theta_f}{(h + t/2 + D/2)^3 H_t} \times 100(\%) \quad (10)$$

Fig. 2. (a) The hysteresis loop of CoFe_2O_4 FNSA with a 26 nm particle size, and (b) the SEM (Scanning Electron Microscopy) image of the CoFe_2O_4 FNSA, (c) the hysteresis loop of CoFe_2O_4 SPNSA with a 7 nm particle size. The inset shows the minor hysteresis loop measured at ± 300 Oe, and (d) the TEM (Transmission Electron Microscopy) image of the CoFe_2O_4 SPNSA.

2.2 Comparison of sensing output performance

Based on the physical model developed in section 2.1, the sensing output performance of an in-vitro GMR biosensor with a single immobilized FNSA or SPNSA is calculated and compared. Figure 3 shows the dependence of sensor width at the fixed sensor geometry (sensor aspect ratio) on the δR and the IF of the in-vitro GMR biosensors. The sensor width, W , of the GMR biosensor is changed from 10 to 80 nm at the different aspect ratio ($L : W$) changed from 3 : 1 to 10 : 1. The purpose of changing the sensor aspect ratio is to explore the effects of vortex magnetization on the surface of FL due to the geometrically-induced demagnetizing factor (Girgis et al., 2000). The H_b , and the h are fixed at a 50 Oe, and a 30 nm, respectively for precise comparison. The CoFe_2O_4 FNSA, and SPNSA has a mean particle size of 26 nm and 7 nm, respectively. The δR and its variation due to the change of W is numerically analyzed by considering both the “effective sensing area”, which is defined as the area formed on the FL surface whose magnetic spins can be coherently

rotated by the stray magnetic field induced by the sensor agent, and the development of “inactive sensing area”, which is not responded by the stray field, due to the increase of IF induced by the geometrically-increased magnetic anisotropy of FL.

Fig. 3. The physical dependence of sensor width, W , on the relative MR, δR , and the interaction factor, IF , (a) δR , GMR biosensor with a FNSA, (b) IF , GMR biosensor with a FNSA, (c) δR , GMR biosensor with a SPNSA, and (d) IF , GMR biosensor with a SPNSA.

As shown in Fig. 3, the in-vitro GMR biosensors with an immobilized single FNSA or SPNSA exhibit the same physical characteristics that the δR is abruptly decreased above the maximized value obtained at the optimized sensor width, W_{op} , and that the IF is almost squarely increased, by increasing the W as well as the aspect ratio. This is supposed to be due to the increase of “inactive sensing area” and the magnetic anisotropy of FL induced by the increased sensor size proportional to the W . However, it is clearly noted that the absolute δR and IF values of the in-vitro GMR biosensor with a FNSA are much larger than those with a SPNSA. As can be clearly seen in Fig. 3-(a) and (c), the δR obtained from the in-vitro GMR biosensor with an aspect ratio of 3: 1 (75 nm (L) $\times$ 25 nm (W_{op})) and an immobilized FNSA is a 2.72 %, while the δR for the in-vitro GMR biosensor with a SPNSA, which has the same aspect ratio (45 nm (L) $\times$ 15 nm (W_{op})), is a 0.013 %. In addition, the variation of IF values depending on the W of the in-vitro GMR biosensor with a FNSA is negligibly small compared to those with a SPNSA as shown in Fig. 3-(b) and (d). The practically allowable sensor size based on the physical limit of current sensor fabrication technology, especially nanoelectronics technology, is another physical parameter to be considered in evaluating the sensing performance. Considering the patterning limit of EBL

(Electron Beam Lithography) technique (> 50 nm) and the geometrically-induced demagnetizing factor of FL directly relevant to the sensor aspect ratio and the IF, the minimum sensor size can be determined in the range between 150 nm (L) $\times$ 50 nm (W) and 250 (nm) $\times$ 50 nm (W). However, as verified in Fig. 3-(a) and (c), the δR values obtained from these sizes of in-vitro GMR biosensor with a SPNSA are too small to be considered for a real biosensor application.

According to the numerically analyzed sensing performance summarized in Fig. 3, it is clearly demonstrated that an in-vitro GMR biosensor with an immobilized single CoFe_2O_4 FNSA is more suitable for SMD due to its higher δR , less IF dependence, and practically allowable sensor size. The large remnant magnetization of single CoFe_2O_4 FNSA allowing to produce a sufficiently large stray field and to maintain extremely small variation of IF is the main physical reason for the technical promise of GMR biosensor with an immobilized single FNSA for SMD.

3. Optimizing the sensor geometry of an in-vitro GMR biosensor with an immobilized FNSA for SMD

In this chapter, the detailed spatial magnetic field interactions between the single CoFe_2O_4 FNSA and the FL of an in-vitro GMR biosensor is numerically analyzed to predict the optimized sensor geometry that maximizes the sensing performance for SMD prior to fabrication. In order to more accurately analyze the spatial magnetic field interactions on the FL surface, the longitudinal and the transverse components of the stray field produced by the FNSA are considered. The optimized sensor geometry at a given remnant magnetic moment of the FNSA is predicted by evaluating the “effective sensing area”. The optimized sensor geometry is expressed in terms of the effective distance (δ), which includes the radius, a , of FNSA, the length of biological entities (especially, DNA including probe), membrane thickness, and the passivation layer, as well as the critical sensor length (l_c), and the critical sensor width (w_c). The experimentally demonstrated sensing performance of an in-vitro GMR biosensor with an immobilized CoFe_2O_4 FNSA is also compared to the numerically calculated sensing performance to confirm the effectiveness of the physical model introduced in this chapter.

3.1 Analytical model for optimizing sensor geometry and geometrical parameters

Figure 4 shows the schematic diagram of an in-vitro GMR biosensor with an immobilized single FNSA (a) and the typical MR curve (b) obtained from the $\text{Si}/\text{Ta}/\text{Ni}_{80}\text{Fe}_{20}/\text{Ir}_{22}\text{Mn}_{78}/\text{Co}_{84}\text{Fe}_{16}/\text{Ru}/\text{Co}_{84}\text{Fe}_{16}/\text{Cu}/\text{Co}_{84}\text{Fe}_{16}/\text{Ni}_{80}\text{Fe}_{20}/\text{Ta}$ exchange biased synthetic GMR spin-valve biosensor. For the numerical calculation, it is assumed that the CoFe_2O_4 FNSA has an $a = 250$ nm, and a mass density of 5.29 g/cm³ (Lee et al., 2007). By considering only the longitudinal field component of the stray field produced by the immobilized single CoFe_2O_4 FNSA, B_x on the surface of FL along the x and y axis from equation (2) is simplified by equation (11) (Schepper et al., 2006).

$$\begin{aligned} B_{x,x\text{-axis}} &= m \cdot \frac{2x^2 - z^2}{(x^2 + z^2)^{5/2}} \\ B_{x,y\text{-axis}} &= m \cdot \frac{-y^2 - z^2}{(y^2 + z^2)^{5/2}} \end{aligned} \quad (11)$$

The calculated magnetic field distribution and the two geometrically critical parameters, which are essential to determine the optimized sensor geometry, are also denoted in Fig. 4-(a). The geometrical parameters of the in-vitro GMR biosensor with an immobilized single FNSA are first determined by considering the longitudinal component of the stray field. The effective magnetization, β , is defined as the ratio of the total magnetization of the CoFe_2O_4 FNSA to the longitudinal field component of the stray field, $\beta = m / B_x$. The δ is defined as $\delta = h + a$. The geometrical parameters, the l_c and the w_c for achieving the optimized sensor geometry, which maximize the sensor output performance, are dependent on β and δ . These geometrical parameters, which determine the "effective sensing area", $l_c \times w_c$, are derived from equation (11) by considering x and y at the points where B_x is equal to the sensor switching field, B_{sw} (with $\beta_{sw} \equiv m / B_x$). The finally determined l_c and w_c are given by,

$$l_c = |2x| = 2\sqrt{2}\delta \sqrt{\frac{\beta_{sw} - \delta^3}{4\beta_{sw} + 5\delta^3}} \quad (12)$$

$$w_c = |2y| = 2\sqrt{\beta_{sw}^{2/3} - \delta^2}$$

The insert in Fig. 4-(b) highlights the two characteristic parameters relevant to the operation of the in-vitro GMR biosensor; the B_{sw} and the detectable field limit (B_{DL}) directly associated with the exchange bias field of the in-vitro GMR biosensor, are defined in terms of the intensity of stray field produced by the single CoFe_2O_4 FNSA. The critical effective distance, δ_c , can be obtained by considering the operating conditions of the GMR biosensor including B_{sw} , B_{DL} , and the M_r of the single CoFe_2O_4 FNSA. If B_x is in the sensor operating range, $B_{sw} < B_x < B_{DL}$, δ can be expressed as a function of β . On the other hand, if B_x is smaller than B_{sw} ($B_{sw} > B_x$), then $l_c = w_c = 0$. Thus, the critical effective distance, δ_c , for the sensor

Fig. 4. (a) A schematic diagram of in-vitro GMR biosensor with an immobilized single FNSA, the field distribution, and the definition of geometrical parameters considering for optimizing sensor geometry, and (b) a typical MR curve of GMR biosensor and the definition of two sensing characteristics parameters.

operation based on the non-switching conditions: $B_{sw} > B_x$ and equation (12) can be determined at $\delta_c = \sqrt[3]{\beta}$. In addition, from equation (12), the aspect ratio, w_c/l_c for the optimized sensor geometry can be expressed by equation (13).

$$w_c/l_c = \frac{2\sqrt{\beta_{sw}^{2/3} - \delta^2}}{2\sqrt{2}\delta \cdot \sqrt{\frac{\beta_{sw} - \delta^3}{4\beta_{sw} + 5\delta^3}}} = \sqrt{\frac{(\beta_{sw}^{2/3} - \delta^2) \cdot (4\beta_{sw} + 5\delta^3)}{2\delta^2(\beta_{sw} - \delta^3)}} \quad (13)$$

The numerically analyzed magnetic field distribution on the surface of the FL finally obtained by equation (13) clearly demonstrates that the optimized geometrical parameters, l_c , and w_c are directly relevant to δ and β_{sw} . In order to more accurately predict the optimized sensor geometry based on the "effective sensing area, $l_c \times w_c$ ", the numerical calculation is extended to two dimensional field component, both longitudinal and transverse field components, on the FL surface. The "Stoner- Wolfarth model" (or the "asteroid curve model") is employed for the detailed calculation (Hirota et al., 2002).

3.2 Optimizing the sensor geometry considering the one dimensional (longitudinal) field component

As described in the analytical model developed in section 3.1, the optimization of sensor geometry with an immobilized CoFe_2O_4 FNSA is based on the determination of l_c and w_c by considering the longitudinal field component of B_x and B_y , on the FL surface. Figure 5 shows the contour diagrams of the magnetic field intensity and its distributions, B_x and B_y on the FL surface as a function of δ (for $\delta = 0.5, 1.0$, and $2.0 \mu\text{m}$). As can be seen in Figs. 5-(a), 5-(c), and 5-(e), the maximum B_x is rapidly decreased from 691.2 to 10.8 G by increasing δ from 0.5 to $2.0 \mu\text{m}$. As shown in Fig. 4-(b), the in-vitro GMR biosensor considered in this model is operated at magnetic field intensity in the range from 12 G (B_{sw}) to 176 G (B_{DL}). Considering these the magnetic characteristics of GMR biosensor, the shaded region observed at $\delta = 0.5 \mu\text{m}$ due to the large field intensity (Fig. 5-(a)) and all the regions shown in Fig. 5-(e) do not contribute to the sensor operation. This indicates that the l_c and the w_c for the optimized sensor geometry based on equation (12) should be determined at $\delta_c < 0.79 \mu\text{m}$, which corresponds to the sensor operating condition of $B_x \leq B_{DL}$. Furthermore, by combining the calculated value of δ_c with the physical parameters of single CoFe_2O_4 FNSA and equation (12), the l_c and the w_c are determined to be $\sim 1.12 \mu\text{m}$, and $\sim 3.52 \mu\text{m}$. Based on the numerical calculation, the aspect ratio (w_c/l_c) for the optimized sensor geometry of in-vitro GMR biosensor with an immobilized single CoFe_2O_4 FNSA ($a = 250 \text{ nm}$) is determined at $w_c/l_c = 3.14$.

The calculation results shown in Fig. 5 clearly demonstrates that the geometrical and systematic design parameters (δ , l_c , and w_c) of the in-vitro GMR biosensor for producing a highly stable sensing performance can be precisely predicted prior to fabrication if the remnant magnetization of the single CoFe_2O_4 FNSA and the GMR characteristics of the sensor are known.

3.3 Optimizing the sensor geometry considering the longitudinal and transverse field components

Dependence of δ on the transverse component, B_y , is also estimated to confirm its physical contribution to the optimization of in-vitro GMR biosensor geometry. Figure 5-(b), 5-(d),

Fig. 5. Calculated contour diagrams of the longitudinal (left column) and transverse (right column) components of the magnetic field produced by an immobilized CoFe_2O_4 FNSA on the FL surface where δ is varied from 0.5 to 2.0 μm . The area defined by the dashed-dotted line and the shaded region show the optimized sensor geometry, and the undetectable region, respectively.

and 5-(f) show the contour diagrams of B_y as a function of δ changed from 0.5 to 2.0 μm . Similar to the calculation results shown in Figs. 5-(a), 5-(c), and 5-(e), the B_y has a strong dependence on δ . However, the distribution of B_y is completely different from B_x . The distribution of B_x on the FL surface shows an ellipsoidal shape with the major axis along the y-axis, while B_y exhibits a distribution that has a maximum and minimum field intensity of $B_{y,\text{max}} \approx 1/3 B_{x,\text{max}}$ at the position of $(\pm\delta, \mp\delta)$. The numerical comparison between B_x and B_y depending on δ suggests that both components of the stray field should be simultaneously

considered for a more accurate prediction of the sensor geometry. Accordingly, the “Stoner-Wolfarth model”: $H_k^{2/3} = H_x^{2/3} + H_y^{2/3}$, is employed to accurately analyze the spatial magnetic field distribution and intensity on the FL surface. Even though the “Stoner-Wolfarth model” assumes that the FL magnetizations are coherently rotated by the stray field and are homogeneous across the entire FL surface, this model is considerably useful in interpreting the physical behavior of the in-vitro GMR biosensor under a highly localized magnetic dipole field from the immobilized single CoFe_2O_4 FNSA. Figure 6 shows the magnetic field distribution and intensity considering both the longitudinal and transverse field components with different effective distances: $\delta = 0.5, 1.0$, and $2.0 \mu\text{m}$. Unlike the ellipsoidal shape of the “effective sensing area” shown in Fig. 5, the coherently rotated magnetization of the FL induced by two-dimensional magnetic field components shows a more complicated and extended “effective sensing area” due to the contribution of the transverse field component. Figure 7 shows the optimized sensor geometry (white line) and the “effective sensing area” (bright gray region) calculated by considering the one-dimensional

Fig. 6. The magnetic field distribution and intensity on the FL surface calculated by considering the longitudinal and transverse field components at the different effective distance of δ . (a) 0.5 , (b) 1.0 , and (c) $2.0 \mu\text{m}$

(Fig. 7-(a)) and the two-dimensional components (Fig. 7-(b)) based on the “Stoner- Wolfarth model”. Although the numerical values of optimized geometrical parameters determined at the effective distance of $\delta = 0.79 \mu\text{m}$ are the same as $l_c = 1.12 \mu\text{m}$, and $w_c = 3.52 \mu\text{m}$, the “effective sensing area” directly relevant to the sensing output performance is completely different. As can be seen in Fig. 7-(b), the “effective sensing area” is extended due to the transverse field component. This correspondingly results in enhancing the output signal of the in-vitro GMR biosensor. However, as can be also seen in Fig. 7-(b), an undetectable area in the vicinity of center of the optimized sensing area is developed due to the spatial magnetic field interaction. Making a GMR biosensor with a larger exchange bias field and introducing a specially designed sensor structure with a high permeability magnetic shield layer are suggested as an effective solution for the undesirable technical problem.

Fig. 7. Comparison of the optimized sensor geometry (square region) and the “effective sensing area” calculated by considering the (a) one-dimensional field component, and (b) two-dimensional field component on the FL surface.

3.4 Demonstration of sensing performance of the in-vitro GMR biosensors with optimized sensor geometry

The sensing performance of an in-vitro GMR biosensor with an immobilized CoFe_2O_4 ferrimagnetic nanobead SA geometrically optimized by the analytical model developed in chapter 3.1 is demonstrated to confirm its practical effectiveness. The CoFe_2O_4 nanobead with a mean radius, a , of 925 nm synthesized by using a modified sol-gel method is considered as a ferrimagnetic nanobead SA. The optimized sensor geometry of the in-vitro GMR biosensor based on the equations (11) ~ (13) as well as considering a 925 nm of mean nanobead size is calculated to determine the “effective sensing area, $l_c \times w_c$ ”. The sensing output performance of the optimized GMR biosensors is evaluated as a function of sensor length, l , at the fixed w_c by controlling the size of CoFe_2O_4 nanobead SA, which is systematically varied in the range of $a = 925 \text{ nm} \pm 20.5 \%$ as shown in Fig. 8-(a).

The controlled nanobead size leads to changing the l at the fixed w_c due to the variation of stray field intensity caused by the change of effective distance. The GMR biosensor used for this demonstration has a structure of $\text{Si}/\text{Ta}(5)/\text{Ni}_{80}\text{Fe}_{20}(2)/\text{Ir}_{22}\text{Mn}_{78}(20)/\text{Co}_{84}\text{Fe}_{16}(2)/\text{Ru}(0.75)/\text{Co}_{84}\text{Fe}_{16}(2)/\text{Cu}(2.3)/\text{Co}_{84}\text{Fe}_{16}(0.5)/\text{Ni}_{80}\text{Fe}_{20}(2.5)/\text{Ta}(3 \text{ nm})$ and is patterned by using an electron beam lithography (EBL) and a typical photolithography. The patterned

Fig. 8. (a) Schematic diagram of in-vitro GMR biosensors with an immobilized CoFe₂O₄ ferrimagnetic nanobead SA with different bead sizes controlled in the range of $a = 925 \text{ nm} \pm 20.5 \%$, (2) the patterned GMR biosensor with the geomtry of $l = 1 \mu\text{m}$, and $w_c = 5 \mu\text{m}$, and (c) GMR behaviour.

GMR biosensor structure and its GMR behaviour for the before and after patterning, and for the hard axis response are shown in Figs. 8(b), and (c), respectively. As can be seen in Fig. 8-(a), the magnetization of FL is orthogonally coupled to the pinned layer, and the stray field produced by the single CoFe₂O₄ nanobead SA is applied to the hard axis of FL magnetization for the detection of output sensing signal.

On the basis of the numerical analysis, the intensity of stray field produced by the CoFe₂O₄ nanobead SA with a radius of 750 (-20.5 %, negative standard deviation), 925 (mean nano bead size), and 1150 nm (+20.5 %, positive standard deviation) is calculated by considering the experimentally obtained M_r values to determine the l_c . The calculated maximum field intensity is a 67.8, 116.5, and 177.1 Oe (G), respectively and the l_c is revealed to be a 0.85, 1.08, and 1.31 μm , respectively at the fixed $w_c = 5 \mu\text{m}$. Figure 9 shows the detected output signal obtained from the in-vitro GMR biosensor shown in Fig. 8-(b). The detected signal is captured by using an oscilloscope. As can be clearly seen in Fig. 9-(a), when a DC magnet with a constant field of 103 Oe (G) is brought proximity to the GMR biosensor, an output signal of $V_{\text{out}} = 6.13 \text{ mV}$ ($V_{\text{output}} = 613 \text{ mV}$ after 100 times amplification using a 741 OP-AMP) is successfully achieved. This is attributed to the MR change of the GMR biosensor, $\Delta R/R_0 = 2.2 \%$, which is exactly equal to the maximum MR ratio obtained along the hard-axis of the

patterned GMR biosensor shown in Fig. 8-(c). At a 103 Oe (G) of field intensity, the “effective sensing area”, $l_c \times w_c$ induced by the DC magnetic field intensity is larger than the patterned sensor geometry of $l = 1 \mu\text{m}$, and $w = 5 \mu\text{m}$. This indicates that all the FL magnetizations are fully rotated by the applied DC magnetic field resulting in exhibiting the maximum MR ratio of 2.2 %. In contrast, the output signals obtained from the in-vitro GMR biosensors activated by the CoFe_2O_4 nanobead SAs show a strong dependence on the size of nanobead SA. As can be seen in Figs. 9-(b), (c), and (d), the output voltage and the $\Delta V_{\text{out}}/V$ of the GMR biosensor activated by the CoFe_2O_4 nanobead SA with a size of 750, 925, and 1150 nm are $V_{\text{out}} = 6.07 \text{ mV}$ ($\Delta V_{\text{out}}/V = 1.2 \%$), $V_{\text{out}} = 6.12 \text{ mV}$ ($\Delta V_{\text{out}}/V = 2.0 \%$), and $V_{\text{out}} = 6.10 \text{ mV}$ ($\Delta V_{\text{out}}/V = 1.7 \%$), respectively. Even though the non-uniformity and the position dependent stray field intensity produced by the nanobead SA can be considered to be partially influenced on the variation of output sensing signal, the observed sensing signal depending on the size of nanobead SA is primarily interpreted in terms of two physical parameters: (a) the change of stray field intensity relevant to the switching field, and (b) the “inactive sensing area” as well as the development of “undetectable sensing area”. As can be

Fig. 9. Output sensing signal captured from the in-vitro GMR biosensor with geometry of $l = 1 \mu\text{m}$ and $w = 5 \mu\text{m}$. (a) activated by DC magnet, (b) activated by 750 nm size CoFe_2O_4 nanobead SA, (c) activated by 925 nm size CoFe_2O_4 nanobead SA, and (d) activated by 1150 nm size CoFe_2O_4 nanobead SA.

confirmed from the calculation results, the $0.85\ \mu\text{m}$ of l_c determined by the $750\ \text{nm}$ size of nanobead SA at the fixed $w_c = 5\ \mu\text{m}$ is smaller than the geometry of patterned GMR biosensor, $l = 1\ \mu\text{m}$ and $w = 5\ \mu\text{m}$, that results in the reduction of MR ratio due to the “inactive sensing area”. In addition, the reduced stray field intensity due to the decrease of nanobead size leads to the reduction of switching field that results in a lower output sensing signal as shown in Fig. 8-(c). The slight decrease of output sensing signal obtained from the GMR biosensor with an $1150\ \text{nm}$ size nanobead SA is thought to be due to the development of “undetectable sensing area” as shown in Fig. 7-(b). The large stray field intensity, around $177.1\ \text{Oe}$ (G), obtained from the large size of nanobead SA is comparable to the exchange bias field of the patterned GMR biosensor. This induces the partial magnetic reversal of pinned layer resulting in a slight MR degradation. Furthermore, the spatial magnetic field interaction due to the large stray field intensity causes to form an undesirable “undetectable sensing area” at the central region of FL surface of the patterned GMR biosensor that leads to the reduction of MR ratio as well as the output sensing signal.

4. Effects of a specially designed magnetic shield layer (MSL) on the sensing performance of an in-vitro GMR biosensor with an immobilized single FNSA for SMD

As previously discussed, the in-vitro GMR biosensor with an immobilized single CoFe_2O_4 FNSA is suitable for SMD. However, according to the numerical analysis on the spatial field interaction of the stray field, which is produced by a single FNSA, on the FL surface, the field distribution is found to be so complicated and non-uniform that it can not be easily interpreted. In particular, the creation of “undershoot field regions”, which are formed at both edges of the maximum field intensity regions as well as the central regions of FL surface resulted from the spatial magnetic field interaction, is revealed as the most severe problem to induce the reduction of output sensing signal and the sensing stability of the in-vitro GMR biosensor. Thus, a new sensor structure, which can solve this technical challenge, is urgently required in a molecular based diagnostic GMR biosensor system for achieving more stable SMD.

In this chapter, a new structure of in-vitro GMR biosensor with a specially designed magnetic shield layer (MSL) is introduced and discussed based on the numerically analyzed calculation results to explore its effectiveness for the improvement of sensing performance as well as the sensing stability. The effects of MSL thickness including magnetic permeability of MSL, and gap width of MSL on the change of “undershoot field” as well as the “stray field intensity” are primarily discussed to demonstrate the physical contribution of MSL to the sensing performance of the in-vitro GMR biosensor considering for SMD.

4.1 Designing of in-vitro GMR biosensor with magnetic shield layer (MSL)

Figure 10-(a) shows a schematic diagram of an in-vitro GMR biosensor with a specially designed MSL. As shown in Fig. 10-(a), the in-vitro GMR biosensor has a single CoFe_2O_4 FNSA immobilized on the FL surface with a distance of $h\ \mu\text{m}$, which is defined as the distance between the FNSA and the FL surface of the GMR biosensor. According to the physical model numerically developed in chapter 3.1 (Schepper et al., 2006), the h can be expressed as equation (14),

$$h = \left(4\pi \times m \frac{1}{B_{DL}}\right)^{1/3} - a \quad (14)$$

$$m = \frac{4\pi}{3} a^3 M_r$$

In this numerical calculation, an IrMn based exchange biased GMR spin-valve device is considered as a sensing element. Thus, to achieve a stable sensing performance, the maximum field intensity produced by the single FNSA should be adjusted to be lower than the exchange bias field of the GMR biosensor. Considering this sensor operating condition, B_{DL} is determined based on the experimentally obtained exchange bias field from the patterned GMR biosensor.

Fig. 10. (a) A schematic diagram of an in-vitro GMR biosensor with a specially designed MSL, and (b) the distribution of x-component of stray field produced by an immobilized single CoFe_2O_4 FNSA, with (solid line) and without (dashed line) MSL

For the numerical calculation based on equation (14), a radius of 250 nm size CoFe_2O_4 FNSA, an 176 G of exchange bias field, and a $0.48 \mu m$ of h are considered. As can be seen in Fig. 10-(a), the calculated h value includes FL/passivation layer (40 nm)/MSL (0 ~ 300 nm)/ Al_2O_3 functional membrane (140 nm; including a length of biological entities such as ten sequence of DNA: 34 nm). The Al_2O_3 functional membrane layer in this sensor structure is used for both maintaining the h depending on the variation of MSL thickness and immobilizing the FNSA using a membrane probe. The high magnetic permeability of materials such as supermalloy and permally are considered as a MSL in this structure. Figure 10-(b) shows a longitudinal field component (x direction) of stray field intensity on the FL surface without (solid line) and with (dashed line) MSL. The distribution of magnetic field intensity shown in Fig. 10-(b) is calculated as a function of distance from the center point of FNSA to the edge of GMR biosensor along the x direction. In addition, B_{max} , B_{U1} , B_{U2} , and l_{eff} are denoted as a maximum magnetic field intensity, an undershoot field, an outer undershoot field, and an effective detectable length, which is the region enabled to be reversed by the stray field, respectively.

4.2 Effects of magnetic shield layer (MSL) on the sensing performance of in-vitro GMR biosensor for SMD

The MSL has a basic geometry of 300 nm thickness, a permeability of supermalloy (5.2×10^5), 1.2 μm gap width, and 8 μm length. To explore the effects of MSL on the sensing performance, the physical parameters of the MSL including its thickness, its gap width, and its permeability are changed from the basic structure. Figure 11 shows the dependence of MSL thickness on the sensing performance compared in terms of the physical sensing parameters such as B_{U1} , B_{max} , I_{eff} , and B_{U2} . As can be seen in Fig. 11-(a), the undesirable “undershoot field region”, B_{U1} , is dramatically reduced when the MSL has a 1 nm of thickness. By further increasing its thickness, B_{U1} is sharply decreased and then it is completely removed above $t_{\text{MSL}} = 100$ nm. This indicates that the MSL needs to have a critical thickness to build up a close magnetic flux between the FNSA and the MSLs allowing for a distinct magnetic shielding effect. Figure 11-(b) shows the dependence of MSL thickness on the change of B_{max} . When the MSL has a 10 nm of thickness and above, the B_{max} is decreased down to 158 G and then it saturates at 153 G by further increasing the MSL thickness above 100 nm. This numerical result along with the increase of I_{eff} shown in Fig. 11-(c) indicates that the effects of MSL on the improvement of sensing performance of the in-vitro GMR biosensor are quite prominent. The reduction of B_{max} from 188 G (above B_{DL}) to 155 G (below the exchange bias field degradation point) and the increase of I_c of the sensing area due to the MSL shielding effects allow the dramatic increase of “effective sensing area” on the FL surface that leads to the increase of sensing output signal of the patterned GMR biosensor. However, as can be seen in Fig. 11-(d), the magnetic dipole field induced in the MSL due to its high magnetic moment generates another undesirable small “undershoot field”, B_{U2} at the vicinity of the MSL edges. Even though its numerical value is small below

Fig. 11. Effects of MSL thickness on the sensing performance of an in-vitro GMR biosensor with an immobilized 250 nm size CoFe_2O_4 FNSA. (a) B_{U1} , (b) B_{max} , (c) I_{eff} , and (d) B_{U2}

12 G, it should be carefully controlled when the MSL is used for especially multi-array sensor architecture. The effects of MSL permeability on the sensing performance is not discussed in details in this chapter as all the physical sensing parameters have the same dependence on the permeability of MSL. However, the numerical calculation results obtained from the GMR biosensor with geometry of a 300 nm MSL thickness, a 1.2 μm of gap width, and a 8 μm of MSL length demonstrates that the MSL with magnetic permeability of at least 100 shows obvious shielding effects. Figure 12 shows the dependence of MSL gap width, which is changed from 0.2 to 2.0 μm , on the sensing performance of an in-vitro GMR biosensor with an immobilized CoFe_2O_4 FNSA. As shown in Fig. 12-(a), the B_{max} is dramatically increased from 150 G to 260 G, which is beyond the sensing limit, B_{DL} , by decreasing the gap width from 0.8 to 0.2 μm . The dramatic increase of magnetic field intensity at the gap width below 0.8 μm is mainly thought to be attributed to the increase of fringe field produced from the gap between two MSLs (This is quite similar to the “fringe field” produced by the head gap from the writer in magnetic recording technology) (Bertram, 1994). The high permeability of two MSLs, which are closely faced each others with small gap, can produce a strong “deep gap bubble field” due to a high magnetic flux density. This leads to increasing the “fringe field” on the surface of MSL gap that would be diverged into the FL surface resulting in the increase of magnetic field intensity on the FL (sensing layer) surface.

As shown in Figs. 12-(b) and 12-(c), the B_{U1} and the l_{eff} are also strongly influenced by the MSL gap width. The B_{U1} is obviously re-developed when the MSL gap width is increased above 1.4 μm . Moreover, l_{eff} is sharply decreased when the MSL gap width is increased above 1.2 μm . The serious degradation of sensing performance relevant to the dramatic

Fig. 12. Effects of MSL gap width on the sensing performance of an in-vitro GMR biosensor with an immobilized 250 nm size CoFe_2O_4 FNSA. (a) B_{U1} , (b) B_{max} , (c) l_{eff} , and (d) B_{U2}

reduction of “effective sensing area” on the sensor surface due to both the re-development of “undershoot field region” and the reduction of critical length, l_c , are mainly attributed to the reduction of shield gap flux density, which is inversely proportional to the MSL gap width. However, considering that the l_c ($\sim 1.56 \mu\text{m}$), which is numerically determined based on the same geometry and configuration of in-vitro GMR biosensor system, is found to be much larger than that without MSL ($\sim 1.12 \mu\text{m}$), it can be readily understood that shielding effects of MSL on the improvement of sensing performance is quite significant.

In summary, it is numerically demonstrated that the MSL is effective to improve the output sensing performance of an in-vitro GMR biosensor with an immobilized FNSA, because it can successfully remove an undesirable “undershoot field region” and enhance the “effective sensing area” due to the increase of l_c . This indicates that an in-vitro GMR biosensor with a specially designed MSL structure can be considered as a promising sensor structure for SMD due to its achievable high sensing signal and stability.

5. Conclusion

The physical sensing characteristics of an in-vitro GMR biosensor with an immobilized single FNSA have been introduced and discussed to explore its feasibility to a single molecular based disease diagnostic biosensor system. According to the theoretically and experimentally analyzed results, the in-vitro GMR biosensor with a FNSA was revealed to be more suitable for SMD than that with a SPNSA due to its higher relative MR, less interaction factor dependence, and practically allowable sensor size. In addition, the analytical models developed in this chapter allowed to readily predicting the optimized sensor geometry of an in-vitro GMR biosensor with a FNSA prior to fabrication if the physical parameters of the FNSA are provided. In particular, the in-vitro GMR biosensor with a specially designed magnetic shield layer (MSL) was demonstrated to be able to effectively control the undesirable “undershoot sensing region” as well as the “undetectable sensing area” on the FL surface. These promising sensing characteristics improved by the newly designed sensor structure allow for achieving both maximized output sensing signal and higher sensor stability that lead to accelerating the more practical applications to the SMD based diagnostic biosensor system.

6. References

- Baselt D. R.; Lee G. U.; Natesan M.; Metzger S. W.; Sheehan P. E. & Colton R. J., (1998). A biosensor based on magnetoresistive technology. *Biosensors and Bioelectronics*, Vol. 13, pp. 731-739, ISSN 0956-5663
- Besse P.; Boero G.; Demierre M.; Pott V. & Popovic R., (2002). Detection of a single magnetic microbead using a miniaturized silicon Hall sensor. *Applied Physics Letters*, Vol. 80, No. 22, pp. 4199-4201, ISSN 0003-6951
- Bertram, N. (1994). *Theory of magnetic recording*, Cambridge, ISBN 0-521-44512-4, New York
- Girgis E.; Schelten J.; Shi J.; Tehrani S. & Goroncin H., (2000). Switching characteristics and magnetization vortices of thin-film cobalt in nanometer-scale patterned arrays. *Applied Physics Letters*, Vol. 76, No. 25, pp. 3780-3782, ISSN 0003-6951
- Graham D. L.; Ferreira H.; Bernardo J.; Freitas P. P. & Cabral J. M. S., (2000). Single magnetic microsphere placement and detection on-chip using current line designs with

- integrated spin valve sensors: Biotechnological applications. *Journal of Applied Physics*, Vol. 91, No. 10, pp. 7786-7788, ISSN 0021-8979
- Graham D. L.; Ferreira H. A.; Freitas P.P. & Cabral J. M. S., (2003). High sensitivity detection of molecular recognition using magnetically labeled biomolecules and magnetoresistive sensors. *Biosensors and Bioelectronics*, Vol. 18, pp. 483-488, ISSN 0956-5663
- Graham D. L.; Ferreira H. A. & Freitas P.P., (2004). Magnetoresistive-based biosensors and biochips. *TRENDS in biotechnology*, Vol. 22, No. 9, pp. 455-462, ISSN 0167-7799
- Hirota E.; Sakakima H. & Inomata K., (2002). *Giant Magnetoresistance Device*, Springer, ISBN 3-540-41819-9, Berlin
- Lagae L.; Wirix-Speethens R.; Das J.; Graham D. L.; Ferreira H.; Freitas P. P.; Borghs G. & Boeck J. De., (2002). On-chip manipulation and magnetization assessment of magnetic bead ensembles by integrated spin-valve sensors. *Journal of Applied Physics*, Vol. 91, No. 10, pp. 7445-7447, ISSN 0021-8979
- Latham A. H.; Tarpara A. N. & Williams M. E., (2007). Magnetic field switching of nanoparticles between orthogonal microfluidic channels, *Analytical Chemistry*, Vol. 79, pp. 5746-5752, ISSN 10.1021
- Lee S.; Bae S.; Takemura Y.; Shim I. B.; Kim T. M.; Kim J.; Lee H. J.; Zurn S. & Kim C., (2007). Self-heating characteristics of cobalt ferrite nanoparticles for hyperthermia application. *Journal of Magnetism and Magnetic Materials*, Vol. 310, pp. 2868-2870, ISSN 0304-8853
- Li G. & Wang S. X., (2003). Analytical and micromagnetic modeling for detection of a single magnetic microbead or nanobead by spin valve sensors. *IEEE Transaction on Magnetics*, Vol. 39, No. 5, pp. 3313-3315, ISSN 10.1109
- Megens M. & Prins M., (2005). Magnetic biochips: a new option for sensitive diagnostics. *Journal of Magnetism and Magnetic Materials*, Vol. 293, pp. 702-708, ISSN 0304-8853
- Miller M. M.; Sheehan P. E.; Edelstein R. L.; Tamanna C. R.; Zhong L.; Bounnak S.; Whiteman L. J. & Colton R. J., (2001). A DNA array sensor utilizing microbeads and magnetoelectronic detection. *Journal of Magnetism and Magnetic Materials*, Vol. 225, pp. 138-144, ISSN 0304-8853
- Ramadan Q.; Samper V.; Poenar D. & Yu C., (2006). Magnetic-based microfluidic platform for biomolecular separation, *Biomedical Microdevices*, Vol. 8, No. 4, pp. 151-158, ISSN 10.1007
- Rife J. C.; Miller M. M.; Sheehan P. E.; Tamanna C. R.; Tondra M. & Whitman L. J., (2003). Design and performance of GMR sensors for the detection of magnetic microbeads in biosensors. *Sensors and Actuators A*, Vol. 107, pp. 209-218, ISSN 0924-4247
- Schepper W.; Schotter J.; Bruckl H. & Reiss G., (2004). Analyzing a magnetic molecule detection system – computer simulation. *Journal of Biotechnology*, Vol. 112, pp. 35-46, ISSN 0168-1656
- Schepper W.; Schotter J.; Bruckl H. & Reiss G., (2006). A magnetic molecule detection system – A comparison of different setups by computer simulation. *Physica B*, Vol. 372, pp. 337-340, ISSN 0921-4526
- Shen W.; Liu X.; Mazumdar D. & Xiao G., (2005). In-situ detection of single micron-sized magnetic beads using magnetic tunnel junction sensors. *Applied Physics Letters*, Vol. 86, pp. 253901-253903, 0003-6951

- Stoner E. C. & Wöhlfarth, (1948). A mechanism of magnetic hysteresis in heterogeneous alloys. *Philosophical Transactions of the Royal Society A*, Vol. 240, pp. 599-642, ISSN 1471-2961
- Tondra M.; Porter M. & Lipert R. J., (2000). Model for detection of immobilized superparamagnetic nanosphere assay labels using giant magnetoresistive sensors. *Journal of Vacuum Science and Technology A*, Vol. 18, No. 4, pp. 1125-1129, ISSN 0734-2101
- Wang S. X.; Bae S. Y.; Li G.; Sun S.; White R. L., Kemp J. T. & Webb C. D., (2005). Towards a magnetic microarray for sensitive diagnostics. *Journal of Magnetism and Magnetic Materials*, Vol. 293, pp. 731-736, ISSN 0304-8853
- William E. M., (2001). *Design and analysis of magnetoresistive recording heads*, John Wiley & Sons, ISBN 0-471-36358-8, New York
- Wirix-Speetjens R.; Fyen W.; Boeck J. D. & Borghs G., (2006). Single magnetic particle detection: Experimental verification of simulated behavior. *Journal of Applied Physics*, Vol. 99, No. 103903, pp. 1-4, ISSN 0021-8979

Intelligent and Biosensors

Edited by Vernon S. Somerset

ISBN 978-953-7619-58-9

Hard cover, 386 pages

Publisher InTech

Published online 01, January, 2010

Published in print edition January, 2010

The use of intelligent sensors have revolutionized the way in which we gather data from the world around us, how we extract useful information from that data, and the manner in which we use the newly obtained information for various operations and decision making. This book is an attempt to highlight the current research in the field of Intelligent and Biosensors, thereby describing state-of-the-art techniques in the field and emerging new technologies, also showcasing some examples and applications.

How to reference

In order to correctly reference this scholarly work, feel free to copy and paste the following:

Seongtae Bae (2010). In-Vitro Magnetoresistive Biosensors for Single Molecular Based Disease Diagnostics: Optimization of Sensor Geometry and Structure, Intelligent and Biosensors, Vernon S. Somerset (Ed.), ISBN: 978-953-7619-58-9, InTech, Available from: <http://www.intechopen.com/books/intelligent-and-biosensors/in-vitro-magnetoresistive-biosensors-for-single-molecular-based-disease-diagnostics-optimization-of->

INTeCH

open science | open minds

InTech Europe

University Campus STeP Ri
Slavka Krautzeka 83/A
51000 Rijeka, Croatia
Phone: +385 (51) 770 447
Fax: +385 (51) 686 166
www.intechopen.com

InTech China

Unit 405, Office Block, Hotel Equatorial Shanghai
No.65, Yan An Road (West), Shanghai, 200040, China
中国上海市延安西路65号上海国际贵都大饭店办公楼405单元
Phone: +86-21-62489820
Fax: +86-21-62489821

© 2010 The Author(s). Licensee IntechOpen. This chapter is distributed under the terms of the [Creative Commons Attribution-NonCommercial-ShareAlike-3.0 License](#), which permits use, distribution and reproduction for non-commercial purposes, provided the original is properly cited and derivative works building on this content are distributed under the same license.